

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135189

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/26/2025	09:25
2	CAL2	1	Water	NA	NA	20.2	7.00	03/26/2025	09:26
3	CAL3	1	Water	NA	NA	20.3	10.02	03/26/2025	09:30
4	ICV	1	Water	NA	NA	20.2	7.01	03/26/2025	09:33
5	CCV1	1	Water	NA	NA	20.3	2.01	03/26/2025	09:35
6	Q1643-06	1	Water	NA	NA	20.9	7.55	03/26/2025	09:45
7	Q1644-03	1	Water	NA	NA	20.5	8.36	03/26/2025	09:55
8	Q1644-03DUP	1	Water	NA	NA	20.6	8.37	03/26/2025	09:56
9	CCV2	1	Water	NA	NA	20.3	12.02	03/26/2025	10:00

WORKLIST(Hardcopy Internal Chain)

135189

WorkList Name : PH W Q1643

WorkList ID : 188545

Department : Wet-Chemistry

Date : 03-26-2025 09:01:04

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1643-06	P A4212	Water	pH	Cool 4 deg C	PSEG03	I31	03/25/2025	9040C
Q1644-03	L 031925	Water	pH	Cool 4 deg C	PSEG03	I31	03/25/2025	9040C

Date/Time 03/26/25 09:15
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 03/26/25 11:40
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]