

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135190

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/26/2025	09:25
2	CAL2	1	Water	NA	NA	20.2	7.00	03/26/2025	09:26
3	CAL3	1	Water	NA	NA	20.3	10.02	03/26/2025	09:30
4	ICV	1	Water	NA	NA	20.2	7.01	03/26/2025	09:33
5	CCV1	1	Water	NA	NA	20.2	2.02	03/26/2025	10:10
6	Q1639-01	1	Solid	20.02	20	24.7	5.63	03/26/2025	10:20
7	Q1639-01DUP	1	Solid	20.03	20	24.8	5.64	03/26/2025	10:21
8	Q1639-02	1	Solid	20.02	20	24.9	5.84	03/26/2025	10:22
9	CCV2	1	Water	NA	NA	20.3	12.02	03/26/2025	10:25

UB 135190

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH Q1639 WorkList ID : 188544 Department : Wet-Chemistry Date : 03-26-2025 09:00:16

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1639-01	TRE-25-0040-TRE-25-0041	Solid	pH	Cool 4 deg C	PSEG03	I31	03/25/2025	9045D
Q1639-02	TRE-25-0042	Solid	pH	Cool 4 deg C	PSEG03	I31	03/25/2025	9045D

Date/Time 03/26/25 09:15
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 03/26/25 11:40
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]