

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135204

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/27/2025	08:20
2	CAL2	1	Water	NA	NA	20.2	7.00	03/27/2025	08:21
3	CAL3	1	Water	NA	NA	20.2	10.02	03/27/2025	08:22
4	ICV	1	Water	NA	NA	20.3	7.02	03/27/2025	08:25
5	CCV1	1	Water	NA	NA	20.3	2.01	03/27/2025	08:30
6	Q1661-01	1	Solid	20.02	20	24.4	6.66	03/27/2025	09:40
7	Q1661-01DUP	1	Solid	20.03	20	24.5	6.67	03/27/2025	09:41
8	CCV2	1	Water	NA	NA	20.3	12.02	03/27/2025	09:50

WORKLIST(Hardcopy Internal Chain)

VB 135204

WorkList Name : ph s q1661

WorkList ID : 188577

Department : Wet-Chemistry

Date : 03-27-2025 08:04:31

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1661-01-B	283	Solid	pH	Cool 4 deg C	PSEG03	I41	03/26/2025	9045D

Date/Time 03/27/25 08:10

Raw Sample Received by: JP Sm

Raw Sample Relinquished by: JP Sm

Date/Time 03/27/25 12:30

Raw Sample Received by: JP Sm

Raw Sample Relinquished by: JP Sm