

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB135212

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	03/27/2025	13:45
2	Q1659-01		1	14	100.00	>212.0	>212.0	03/27/2025	14:15
3	Q1659-01DUP		1	14	100.00	>212.0	>212.0	03/27/2025	14:45
4	Q1659-02		1	13	100.00	>212.0	>212.0	03/27/2025	15:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66135212

WorkList Name : FP-3-26-

WorkList ID : 188570

Department : Wet-Chemistry

Date : 03-26-2025 16:31:23

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1659-01 F	3808	Water	Flash Point	Cool 4 deg C	PSEG03	I31	03/26/2025	1010B
Q1659-02 F	3809	Water	Flash Point	Cool 4 deg C	PSEG03	I31	03/26/2025	1010B

Date/Time 03/27/2025 13:35
 Raw Sample Received by: RM CWG
 Raw Sample Relinquished by: J. Leluc

Date/Time 03/27/2025 15:50
 Raw Sample Received by: M. Coot
 Raw Sample Relinquished by: RM Leluc