

LB135219


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

48162



A= 0.0000e+000

B= 2.0641e-004

C= 3.1020e-002

Rho= 0.9998855

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% RD
0.0	0.000	0.030	0.030	-3	0.000	-3					
0.2	0.200	0.229	0.029	959	0.0 %	959					
2.5	2.500	2.427	-0.073	11608	0.0 %	11608					
5.0	5.000	4.950	-0.050	23830	0.0 %	23830					
7.5	7.500	7.591	0.091	36627	0.0 %	36627					
10.0	10.000	9.972	-0.028	48162	0.0 %	48162					

4/10
-3
15
-3
-1
1
0

LB135219 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 27 Mar 2025 14:28:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-3	-	0.0000	7471B	PPB	27 Mar 2025 14:49:45	S	Std
0.2 - 1	50.2	959	-	0.2000	7471B	PPB	27 Mar 2025 14:52:02	S	Std
2.5 - 1	52.5	11608	-	2.5000	7471B	PPB	27 Mar 2025 14:54:20	S	Std
5.0 - 1	55	23830	-	5.0000	7471B	PPB	27 Mar 2025 14:56:37	S	Std
7.5 - 1	57.5	36627	-	7.5000	7471B	PPB	27 Mar 2025 15:01:22	S	Std
10.0 - 1	57.5	48162	-	10.0000	7471B	PPB	27 Mar 2025 15:03:37	S	Std
ICV64 - 1	ICV64	18176	3.7828	-	7471B	PPB	27 Mar 2025 15:06:35	U	SMPL
ICB64 - 1	ICB64	-334	-0.0379	-	7471B	PPB	27 Mar 2025 15:08:51	U	SMPL
CCV91 - 1	CCV91	24021	4.9893	-	7471B	PPB	27 Mar 2025 15:11:09	U	SMPL
CCB91 - 1	CCB91	-411	-0.0538	-	7471B	PPB	27 Mar 2025 15:13:25	U	SMPL
CRA - 1	CRA	941	0.2253	-	7471B	PPB	27 Mar 2025 15:15:43	U	SMPL
HighStd - 1	HighStd	46283	9.5845	-	7471B	PPB	27 Mar 2025 15:17:59	U	SMPL
ChkStd - 1	ChkStd	32218	6.6813	-	7471B	PPB	27 Mar 2025 15:22:59	U	SMPL
PB167337BL - 1	PBS	-546	-0.0817	-	7471B	PPB	27 Mar 2025 15:25:16	U	SMPL
PB167337BS - 1	LCSS	18493	3.8482	-	7471B	PPB	27 Mar 2025 15:27:37	U	SMPL
Q1642-01 - 1	PAT-1-032525	264803	54.6901	-	7471B	PPB	27 Mar 2025 15:29:53	U	SMPL
Q1643-03 - 1	030825-COMP	5432	1.1523	-	7471B	PPB	27 Mar 2025 15:32:12	U	SMPL
Q1644-02 - 1	12325-C	-219	-0.0142	-	7471B	PPB	27 Mar 2025 15:35:38	U	SMPL
Q1645-01 - 1	TP-4	500	0.1342	-	7471B	PPB	27 Mar 2025 15:37:57	U	SMPL
Q1645-05 - 1	TP-1	760	0.1879	-	7471B	PPB	27 Mar 2025 15:40:13	U	SMPL
CCV92 - 1	CCV92	23553	4.8927	-	7471B	PPB	27 Mar 2025 15:42:28	U	SMPL
CCB92 - 1	CCB92	-271	-0.0249	-	7471B	PPB	27 Mar 2025 15:44:43	U	SMPL
Q1645-05DUP - 1	TP-1DUP	810	0.1982	-	7471B	PPB	27 Mar 2025 15:47:01	U	SMPL
Q1645-05MS - 1	TP-1MS	17577	3.6592	-	7471B	PPB	27 Mar 2025 15:49:16	U	SMPL
Q1645-05MSD - 1	TP-1MSD	16364	3.4088	-	7471B	PPB	27 Mar 2025 15:51:32	U	SMPL
Q1647-01 - 1	HR-SPILL-DEBRIS	-312	-0.0334	-	7471B	PPB	27 Mar 2025 15:53:54	U	SMPL
Q1648-01 - 1	TP-4	11324	2.3685	-	7471B	PPB	27 Mar 2025 15:56:13	U	SMPL
Q1648-03 - 1	TP-5	31228	6.4769	-	7471B	PPB	27 Mar 2025 15:58:29	U	SMPL
Q1648-05 - 1	TP-3	5056	1.0746	-	7471B	PPB	27 Mar 2025 16:00:52	U	SMPL
Q1648-07 - 1	TP-6	291	0.0911	-	7471B	PPB	27 Mar 2025 16:03:33	U	SMPL
Q1648-09 - 1	TP-7	15384	3.2065	-	7471B	PPB	27 Mar 2025 16:05:52	U	SMPL
Q1648-11 - 1	TP-8	2014	0.4467	-	7471B	PPB	27 Mar 2025 16:08:07	U	SMPL
CCV93 - 1	CCV93	24078	5.0011	-	7471B	PPB	27 Mar 2025 16:10:26	U	SMPL
CCB93 - 1	CCB93	-348	-0.0408	-	7471B	PPB	27 Mar 2025 16:12:43	U	SMPL
Q1648-13 - 1	TP-2	9443	1.9802	-	7471B	PPB	27 Mar 2025 16:15:02	U	SMPL
Q1650-01 - 1	STOCK-PILE-BIN2	13830	2.8857	-	7471B	PPB	27 Mar 2025 16:17:17	U	SMPL
Q1654-01 - 1	RT3407	3650	0.7844	-	7471B	PPB	27 Mar 2025 16:19:36	U	SMPL
Q1656-01 - 1	VNJ239	3200	0.6915	-	7471B	PPB	27 Mar 2025 16:21:58	U	SMPL
Q1657-01 - 1	72-11998	3170	0.6854	-	7471B	PPB	27 Mar 2025 16:24:15	U	SMPL
Q1661-02 - 1	351	876	0.2118	-	7471B	PPB	27 Mar 2025 16:26:33	U	SMPL
Q1662-01 - 1	TP-6	862	0.2089	-	7471B	PPB	27 Mar 2025 16:28:51	U	SMPL
Q1662-05 - 1	TP-7	747	0.1852	-	7471B	PPB	27 Mar 2025 16:31:08	U	SMPL
Q1645-05LX5 - 1		74	0.0463	-	7471B	PPB	27 Mar 2025 16:33:25	U	SMPL
Q1645-05A - 1		18405	3.8301	-	7471B	PPB	27 Mar 2025 16:35:41	U	SMPL
CCV94 - 1	CCV94	23743	4.9319	-	7471B	PPB	27 Mar 2025 16:37:57	U	SMPL
CCB94 - 1	CCB94	-414	-0.0544	-	7471B	PPB	27 Mar 2025 16:40:15	U	SMPL