

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB135220

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	03/28/2025	08:20
2	Q1655-01		1	16	100.00	>212.0	>212.0	03/28/2025	08:50
3	Q1655-01DUP		1	16	100.00	>212.0	>212.0	03/28/2025	09:20
4	Q1663-01		1	13	100.00	>212.0	>212.0	03/28/2025	09:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6135220

WorkList Name : FP-3-27+ WorkList ID : 188601 Department : Wet-Chemistry Date : 03-27-2025 15:18:15

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1655-01 C	TANK-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	I31	03/26/2025	1010B
Q1663-01	0324-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	I41	03/27/2025	1010B

Date/Time 03/28/2025 08:10
Raw Sample Received by: RM (w/1)
Raw Sample Relinquished by: JLC/CI

Date/Time 03/28/2025 10:30
Raw Sample Received by: JLC/CI
Raw Sample Relinquished by: RM (w/1)