

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB135240

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/31/2025	13:15
2	CAL2	1	Water	NA	NA	20.2	7.01	03/31/2025	13:11
3	CAL3	1	Water	NA	NA	20.3	10.02	03/31/2025	13:15
4	ICV	1	Water	NA	NA	20.3	6.99	03/31/2025	13:22
5	CCV1	1	Water	NA	NA	20.1	2.01	03/31/2025	13:25
6	Q1663-02	1	Water	NA	NA	20.1	7.31	03/31/2025	13:35
7	Q1663-02DUP	1	Water	NA	NA	20.2	7.32	03/31/2025	13:36
8	Q1663-03	1	Water	NA	NA	22.1	7.39	03/31/2025	13:40
9	CCV2	1	Water	NA	NA	20.3	12.02	03/31/2025	13:44

WORKLIST(Hardcopy Internal Chain)

13135240

WorkList Name : PH Q1663 W

WorkList ID : 188642

Department : Wet-Chemistry

Date : 03-31-2025 12:50:23

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1663-02	0324-A	Water	pH	Cool 4 deg C	PSEG03	I41	03/27/2025	9040C
Q1663-03	0324-B	Water	pH	Cool 4 deg C	PSEG03	I41	03/27/2025	9040C

Date/Time 03/31/25 13:10

Raw Sample Received by: RLWC

Raw Sample Relinquished by: L.C. sun

Date/Time 03/31/25 16:00

Raw Sample Received by: L.C. sun

Raw Sample Relinquished by: RLWC