

LB185246


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

44271



A= 0.0000e+000

B= 2.2723e-004

C= 5.5230e-002

Rho= 0.9996299

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/00
0.0	0.000	0.058	0.058	10	0.000	10	0				15
0.2	0.200	0.230	0.030	769	0.0 %	769					4
2.5	2.500	2.408	-0.092	10352	0.0 %	10352					1
5.0	5.000	5.064	0.064	22043	0.0 %	22043					2
7.5	7.500	7.326	-0.174	31996	0.0 %	31996					1
10.0	10.000	10.115	0.115	44271	0.0 %	44271					

LB135246 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 Apr 2025 10:18:51

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	10	-	0.0000	7470A	PPB	01 Apr 2025 10:24:38	S	Std
0.2 - 1	50.2	769	-	0.2000	7470A	PPB	01 Apr 2025 10:31:18	S	Std
2.5 - 1	52.5	10352	-	2.5000	7470A	PPB	01 Apr 2025 10:33:35	S	Std
5.0 - 1	55	22043	-	5.0000	7470A	PPB	01 Apr 2025 10:35:52	S	Std
7.5 - 1	57.5	31996	-	7.5000	7470A	PPB	01 Apr 2025 10:41:38	S	Std
10.0 - 1	510	44271	-	10.0000	7470A	PPB	01 Apr 2025 10:43:54	S	Std
ICV68 - 1	ICV68	17004	3.9191	-	7470A	PPB	01 Apr 2025 10:51:56	U	SMPL
ICB68 - 1	ICB68	133	0.0855	-	7470A	PPB	01 Apr 2025 10:56:51	U	SMPL
CCV07 - 1	CCV07	21714	4.9894	-	7470A	PPB	01 Apr 2025 10:59:05	U	SMPL
CCB07 - 1	CCB07	-85	0.0359	-	7470A	PPB	01 Apr 2025 11:01:20	U	SMPL
CRA - 1	CRA	799	0.2368	-	7470A	PPB	01 Apr 2025 11:03:37	U	SMPL
HighStd - 1	HighStd	43266	9.8867	-	7470A	PPB	01 Apr 2025 11:05:52	U	SMPL
ChkStd - 1	ChkStd	30334	6.9481	-	7470A	PPB	01 Apr 2025 11:11:06	U	SMPL
PB167414BL - 1	PBW	-487	-0.0554	-	7470A	PPB	01 Apr 2025 11:13:22	U	SMPL
PB167414BS - 1	LCSW	17855	4.1125	-	7470A	PPB	01 Apr 2025 11:15:45	U	SMPL
Q1671-04 - 1	WC-1	-250	-0.0016	-	7470A	PPB	01 Apr 2025 11:18:01	U	SMPL
Q1672-04 - 1	TP-8	-36	0.0470	-	7470A	PPB	01 Apr 2025 11:20:19	U	SMPL
Q1674-02 - 1	RT5358	-13	0.0523	-	7470A	PPB	01 Apr 2025 11:22:36	U	SMPL
Q1674-04 - 1	72-11991	36	0.0634	-	7470A	PPB	01 Apr 2025 11:24:53	U	SMPL
Q1679-04 - 1	TP-10	27	0.0614	-	7470A	PPB	01 Apr 2025 11:27:08	U	SMPL
CCV08 - 1	CCV08	22079	5.0723	-	7470A	PPB	01 Apr 2025 11:29:24	U	SMPL
CCB08 - 1	CCB08	-39	0.0464	-	7470A	PPB	01 Apr 2025 11:31:39	U	SMPL
Q1679-08 - 1	TP-9	-112	0.0298	-	7470A	PPB	01 Apr 2025 11:33:56	U	SMPL
Q1679-08DUP - 1	TP-9DUP	-59	0.0418	-	7470A	PPB	01 Apr 2025 11:36:11	U	SMPL
Q1679-08MS - 1	TP-9MS	15831	3.6526	-	7470A	PPB	01 Apr 2025 11:38:27	U	SMPL
Q1679-08MSD - 1	TP-9MSD	15374	3.5487	-	7470A	PPB	01 Apr 2025 11:40:42	U	SMPL
PB167415BL - 1	PBW	-232	0.0025	-	7470A	PPB	01 Apr 2025 11:43:00	U	SMPL
PB167415BS - 1	LCSW	16256	3.7491	-	7470A	PPB	01 Apr 2025 11:45:17	U	SMPL
Q1664-04 - 1	P001-BBDGA-001-01	87	0.0750	-	7470A	PPB	01 Apr 2025 11:52:28	U	SMPL
Q1664-04DUP - 1	P001-BBDGA-001-01DUP	90	0.0757	-	7470A	PPB	01 Apr 2025 11:54:46	U	SMPL
Q1664-05 - 1	Q1664-04MS	15637	3.6085	-	7470A	PPB	01 Apr 2025 11:57:03	U	SMPL
Q1664-06 - 1	Q1664-04MSD	15060	3.4774	-	7470A	PPB	01 Apr 2025 11:59:19	U	SMPL
CCV09 - 1	CCV09	20921	4.8092	-	7470A	PPB	01 Apr 2025 12:01:36	U	SMPL
CCB09 - 1	CCB09	46	0.0657	-	7470A	PPB	01 Apr 2025 12:03:54	U	SMPL
Q1664-08 - 1	P001-BBDGA-001-02	-57	0.0423	-	7470A	PPB	01 Apr 2025 12:06:12	U	SMPL
Q1664-10 - 1	P001-BBDGA-002-01	3	0.0559	-	7470A	PPB	01 Apr 2025 12:08:28	U	SMPL
Q1664-12 - 1	P001-BBDGA-003-01	49	0.0664	-	7470A	PPB	01 Apr 2025 12:10:43	U	SMPL
Q1664-14 - 1	P001-BBDGA-004-01	41	0.0645	-	7470A	PPB	01 Apr 2025 12:12:59	U	SMPL
Q1664-16 - 1	P001-BBDGA-005-01	6	0.0566	-	7470A	PPB	01 Apr 2025 12:15:15	U	SMPL
Q1664-18 - 1	P001-BBDGA-006-01	5	0.0564	-	7470A	PPB	01 Apr 2025 12:17:32	U	SMPL
Q1664-20 - 1	P001-BBDGA-007-01	-69	0.0396	-	7470A	PPB	01 Apr 2025 12:19:48	U	SMPL
Q1664-22 - 1	P001-BBDGA-008-01	29	0.0618	-	7470A	PPB	01 Apr 2025 12:22:05	U	SMPL
CCV10 - 1	CCV10	20186	4.6422	-	7470A	PPB	01 Apr 2025 12:24:22	U	SMPL
CCB10 - 1	CCB10	-29	0.0486	-	7470A	PPB	01 Apr 2025 12:26:38	U	SMPL
PB167372TB - 1	PB167372TB	-30	0.0484	-	7470A	PPB	01 Apr 2025 12:28:55	U	SMPL
PB167373TB - 1	PB167373TB	13	0.0582	-	7470A	PPB	01 Apr 2025 12:31:11	U	SMPL
Q1679-08LX5 - 1		27	0.0614	-	7470A	PPB	01 Apr 2025 12:33:26	U	SMPL
Q1679-08A - 1		15134	3.4942	-	7470A	PPB	01 Apr 2025 12:35:41	U	SMPL
Q1664-04LX5 - 1		-72	0.0389	-	7470A	PPB	01 Apr 2025 12:37:57	U	SMPL
Q1664-04A - 1		14862	3.4324	-	7470A	PPB	01 Apr 2025 13:00:35	U	SMPL
CCV11 - 1	CCV11	21170	4.8658	-	7470A	PPB	01 Apr 2025 13:02:50	U	SMPL
CCB11 - 1	CCB11	-35	0.0473	-	7470A	PPB	01 Apr 2025 13:05:07	U	SMPL