

LB135264


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

48847



A= 0.0000e+000

B= 2.0763e-004

C= 7.2572e-002

Rho= 0.9991130

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.064	0.064	-40	0.000	-40					
0.2	0.200	0.237	0.037	793	0.0 %	793					
2.5	2.500	2.483	-0.017	11609	0.0 %	11609					
5.0	5.000	5.005	0.005	23757	0.0 %	23757					
7.5	7.500	7.195	-0.305	34304	0.0 %	34304					
10.0	10.000	10.215	0.215	48847	0.0 %	48847					

19
11
0
4
2

LB135264 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 02 Apr 2025 10:29:02

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	LSO 0	-40	-	0.0000	7471B	PPB	02 Apr 2025 10:42:42	S	Std
0.2 - 1	50.2	793	-	0.2000	7471B	PPB	02 Apr 2025 10:44:59	S	Std
2.5 - 1	52.5	11609	-	2.5000	7471B	PPB	02 Apr 2025 10:47:17	S	Std
5.0 - 1	55	23757	-	5.0000	7471B	PPB	02 Apr 2025 10:49:33	S	Std
7.5 - 1	57.5	34304	-	7.5000	7471B	PPB	02 Apr 2025 10:51:52	S	Std
10.0 - 1	510	48847	-	10.0000	7471B	PPB	02 Apr 2025 10:54:11	S	Std
ICV71 - 1	ICV71	17063	3.6154	-7471B	PPB	PPB	02 Apr 2025 11:02:41	U	SMPL
ICB71 - 1	ICB71	-298	0.0107	-7471B	PPB	PPB	02 Apr 2025 11:04:58	U	SMPL
CCV18 - 1	CCV18	23669	4.9871	-7471B	PPB	PPB	02 Apr 2025 11:07:16	U	SMPL
CCB18 - 1	CCB18	-409	-0.0124	-7471B	PPB	PPB	02 Apr 2025 11:09:31	U	SMPL
CRA - 1	CRA	751	0.2285	-7471B	PPB	PPB	02 Apr 2025 11:11:49	U	SMPL
HighStd - 1	HighStd	47889	10.0160	-7471B	PPB	PPB	02 Apr 2025 11:14:04	U	SMPL
ChkStd - 1	ChkStd	33017	6.9280	-7471B	PPB	PPB	02 Apr 2025 11:19:08	U	SMPL
PB167423BL - 1	PBS	-506	-0.0325	-7471B	PPB	PPB	02 Apr 2025 11:21:25	U	SMPL
PB167423BS - 1	LCSS	15792	3.3515	-7471B	PPB	PPB	02 Apr 2025 11:26:33	U	SMPL
Q1681-01 - 1	TP-12	677	0.2131	-7471B	PPB	PPB	02 Apr 2025 11:28:49	U	SMPL
Q1684-01 - 1	SU-04-040125	1311	0.3448	-7471B	PPB	PPB	02 Apr 2025 11:36:33	U	SMPL
Q1684-01DUP - 1	SU-04-040125DUP	1072	0.2952	-7471B	PPB	PPB	02 Apr 2025 11:38:49	U	SMPL
Q1684-01MS - 1	SU-04-040125MS	21868	4.6131	-7471B	PPB	PPB	02 Apr 2025 11:43:43	U	SMPL
Q1684-01MSD - 1	SU-04-040125MSD	19865	4.1972	-7471B	PPB	PPB	02 Apr 2025 11:45:59	U	SMPL
CCV19 - 1	CCV19	22738	4.7938	-7471B	PPB	PPB	02 Apr 2025 11:48:20	U	SMPL
CCB19 - 1	CCB19	-416	-0.0138	-7471B	PPB	PPB	02 Apr 2025 11:50:38	U	SMPL
Q1685-01 - 1	HR-02-040125	1274	0.3371	-7471B	PPB	PPB	02 Apr 2025 11:52:57	U	SMPL
Q1686-01 - 1	EO-03-040125	471	0.1704	-7471B	PPB	PPB	02 Apr 2025 11:55:12	U	SMPL
Q1687-01 - 1	72-12016	1614	0.4077	-7471B	PPB	PPB	02 Apr 2025 11:57:29	U	SMPL
Q1688-01 - 1	OR-02-040125	1502	0.3844	-7471B	PPB	PPB	02 Apr 2025 11:59:45	U	SMPL
Q1692-01 - 1	NB-08-04012025	1181	0.3178	-7471B	PPB	PPB	02 Apr 2025 12:02:00	U	SMPL
Q1693-01 - 1	WCS-TP-3	2382	0.5672	-7471B	PPB	PPB	02 Apr 2025 12:04:16	U	SMPL
Q1693-05 - 1	WCS-TP-2	5581	1.2314	-7471B	PPB	PPB	02 Apr 2025 12:06:33	U	SMPL
Q1693-09 - 1	WCS-TP-1	1559	0.3963	-7471B	PPB	PPB	02 Apr 2025 12:08:51	U	SMPL
Q1684-01LX5 - 1		242	0.1228	-7471B	PPB	PPB	02 Apr 2025 12:11:09	U	SMPL
Q1684-01A - 1		20760	4.3831	-7471B	PPB	PPB	02 Apr 2025 12:13:25	U	SMPL
CCV20 - 1	CCV20	22675	4.7807	-7471B	PPB	PPB	02 Apr 2025 12:15:41	U	SMPL
CCB20 - 1	CCB20	-345	0.0009	-7471B	PPB	PPB	02 Apr 2025 12:18:00	U	SMPL