

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB135265

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	04/03/2025	13:10
2	Q1683-01		1	14	100.00	>212.0	>212.0	04/03/2025	13:40
3	Q1683-01DUP		1	14	100.00	>212.0	>212.0	04/03/2025	14:10
4	Q1683-02		1	16	100.00	>212.0	>212.0	04/03/2025	14:40
5	Q1683-03		1	13	100.00	>212.0	>212.0	04/03/2025	15:10
6	Q1700-01		1	14	100.00	>212.0	>212.0	04/03/2025	15:40
7	Q1700-02		1	15	100.00	>212.0	>212.0	04/03/2025	16:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6135265

WorkList Name : FP-4-01

WorkList ID : 188668

Department : Wet-Chemistry

Date : 04-01-2025 15:44:56

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1683-01 C	60432	Water	Flash Point	Cool 4 deg C	PSEG03	I41	04/01/2025	1010B
Q1683-02 C	60437	Water	Flash Point	Cool 4 deg C	PSEG03	I41	04/01/2025	1010B
Q1683-03 C	60438	Water	Flash Point	Cool 4 deg C	PSEG03	I41	04/01/2025	1010B
Q1700-01 C	32725	Water	Flash Point	Cool 4 deg C	PSEG03	I31	04/02/2025	1010B
Q1700-02 C	32525	Water	Flash Point	Cool 4 deg C	PSEG03	I31	04/02/2025	1010B

Date/Time 04/03/2025 13:00
 Raw Sample Received by: RM (wg)
 Raw Sample Relinquished by: RM (wg)

Date/Time 04/03/2025 16:50
 Raw Sample Received by: RM (wg)
 Raw Sample Relinquished by: RM (wg)