

LB135269


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39587



A= 0.0000e+000

B= 2.5692e-004

C= 2.8351e-002

Rho= 0.9989419

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.046	0.046	68	0.000	68	0				
0.2	0.200	0.233	0.033	798	0.0 %	798					
2.5	2.500	2.466	-0.034	9490	0.0 %	9490					
5.0	5.000	5.100	0.100	19741	0.0 %	19741					
7.5	7.500	7.155	-0.345	27740	0.0 %	27740					
10.0	10.000	10.199	0.199	39587	0.0 %	39587					

17
-1
2
-5
2

LB135269 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Apr 2025 13:29:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	68	-	0.0000	7470A	PPB	02 Apr 2025 13:40:28	S	Std
0.2 - 1	50.2	798	-	0.2000	7470A	PPB	02 Apr 2025 13:42:45	S	Std
2.5 - 1	52.5	9490	-	2.5000	7470A	PPB	02 Apr 2025 13:45:02	S	Std
5.0 - 1	55	19741	-	5.0000	7470A	PPB	02 Apr 2025 13:47:19	S	Std
7.5 - 1	57.5	27740	-	7.5000	7470A	PPB	02 Apr 2025 13:55:02	S	Std
10.0 - 1	510	39587	-	10.0000	7470A	PPB	02 Apr 2025 14:00:04	S	Std
ICV73 - 1	ICV73	15602	4.0368	-	7470A	PPB	02 Apr 2025 14:10:13	U	SMPL
ICB73 - 1	ICB73	-215	-0.0269	-	7470A	PPB	02 Apr 2025 14:12:29	U	SMPL
CCV23 - 1	CCV23	18496	4.7803	-	7470A	PPB	02 Apr 2025 14:14:46	U	SMPL
CCB23 - 1	CCB23	-229	-0.0305	-	7470A	PPB	02 Apr 2025 14:17:01	U	SMPL
CRA - 1	CRA	676	0.2020	-	7470A	PPB	02 Apr 2025 14:19:18	U	SMPL
HighStd - 1	HighStd	37584	9.6843	-	7470A	PPB	02 Apr 2025 14:21:33	U	SMPL
ChkStd - 1	ChkStd	23255	6.0030	-	7470A	PPB	02 Apr 2025 14:23:49	U	SMPL
PB167425BL - 1	PBW	-362	-0.0647	-	7470A	PPB	02 Apr 2025 14:26:07	U	SMPL
Q1666-01 - 1	TW-WTS-05	-351	-0.0618	-	7470A	PPB	02 Apr 2025 14:30:42	U	SMPL
Q1690-01 - 1	MW112010	94	0.0525	-	7470A	PPB	02 Apr 2025 14:33:00	U	SMPL
Q1690-01DUP - 1	MW112010DUP	-10	0.0258	-	7470A	PPB	02 Apr 2025 14:35:17	U	SMPL
Q1690-01MS - 1	MW112010MS	16137	4.1742	-	7470A	PPB	02 Apr 2025 14:40:05	U	SMPL
Q1690-01MSD - 1	MW112010MSD	15761	4.0776	-	7470A	PPB	02 Apr 2025 14:42:19	U	SMPL
CCV24 - 1	CCV24	18250	4.7171	-	7470A	PPB	02 Apr 2025 14:44:37	U	SMPL
CCB24 - 1	CCB24	-258	-0.0379	-	7470A	PPB	02 Apr 2025 14:46:55	U	SMPL
Q1690-01LX5 - 1		-5	0.0271	-	7470A	PPB	02 Apr 2025 14:54:11	U	SMPL
Q1690-01A - 1		14334	3.7110	-	7470A	PPB	02 Apr 2025 14:56:26	U	SMPL
PB167425BS - 1	LCSW	13207	3.4215	-	7470A	PPB	02 Apr 2025 14:58:41	U	SMPL
CCV25 - 1	CCV25	17494	4.5229	-	7470A	PPB	02 Apr 2025 15:03:33	U	SMPL
CCB25 - 1	CCB25	-281	-0.0438	-	7470A	PPB	02 Apr 2025 15:05:50	U	SMPL