

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB135276

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/02/2025	15:30
2	CAL2	1	Water	NA	NA	20.2	7.00	04/02/2025	15:31
3	CAL3	1	Water	NA	NA	20.3	10.02	04/02/2025	15:35
4	ICV	1	Water	NA	NA	20.2	7.01	04/02/2025	15:38
5	CCV1	1	Water	NA	NA	20.3	2.01	04/02/2025	15:44
6	Q1694-04	1	Solid	20.02	20	22.6	7.74	04/02/2025	15:50
7	Q1694-04DUP	1	Solid	20.03	20	22.7	7.75	04/02/2025	15:51
8	Q1695-04	1	Solid	20.04	20	22.8	9.67	04/02/2025	16:10
9	Q1695-08	1	Solid	20.02	20	22.6	8.70	04/02/2025	16:20
10	Q1698-02	1	Solid	20.05	20	24.6	10.55	04/02/2025	16:30
11	Q1698-08	1	Solid	20.04	20	23.7	10.57	04/02/2025	16:35
12	CCV2	1	Water	NA	NA	20.2	12.02	04/02/2025	16:40

WORKLIST(Hardcopy Internal Chain)

17135276

WorkList Name : corrosivity q1695 WorkList ID : 188684 Department : Wet-Chemistry Date : 04-02-2025 14:38:11

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1694-04	TP-18	Solid	Corrosivity	Cool 4 deg C	PSEG03	I41	04/01/2025	9045D
Q1695-04	TT-2	Solid	Corrosivity	Cool 4 deg C	PSEG03	I41	04/02/2025	9045D
Q1695-08	TT-3	Solid	Corrosivity	Cool 4 deg C	PSEG03	I41	04/02/2025	9045D
Q1698-02	B-9	Solid	Corrosivity	Cool 4 deg C	TULL02	I41	04/02/2025	9045D
Q1698-08	B-10	Solid	Corrosivity	Cool 4 deg C	TULL02	I41	04/02/2025	9045D

Date/Time 04/02/25 14:55

Raw Sample Received by: JB CWC

Raw Sample Relinquished by: JB CWC

Date/Time 04/02/25

Raw Sample Received by: JB CWC

Raw Sample Relinquished by: JB CWC

18.00

JB CWC