

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB135278

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/03/2025	08:15
2	CAL2	1	Water	NA	NA	20.3	7.01	04/03/2025	08:16
3	CAL3	1	Water	NA	NA	20.3	10.02	04/03/2025	08:20
4	ICV	1	Water	NA	NA	20.3	7.03	04/03/2025	08:25
5	CCV1	1	Water	NA	NA	20.2	2.01	04/03/2025	08:30
6	Q1683-01	1	Water	NA	NA	21.8	4.53	04/03/2025	08:40
7	Q1683-02	1	Water	NA	NA	22.2	5.56	04/03/2025	08:45
8	Q1683-03	1	Water	NA	NA	22.5	5.57	04/03/2025	08:50
9	Q1700-01	1	Water	NA	NA	20.4	5.29	04/03/2025	09:00
10	Q1700-02	1	Water	NA	NA	20.7	5.35	04/03/2025	09:05
11	Q1702-03	1	Water	NA	NA	20.7	6.22	04/03/2025	09:11
12	Q1710-01	1	Water	NA	NA	20.3	7.67	04/03/2025	09:20
13	Q1710-01DUP	1	Water	NA	NA	20.4	7.68	04/03/2025	09:21
14	CCV2	1	Water	NA	NA	20.3	12.02	04/03/2025	09:25

WORKLIST(Hardcopy Internal Chain)

LB135278

WorkList Name : PH Q1981

WorkList ID : 188688

Department : Wet-Chemistry

Date : 04-03-2025 07:53:00

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1683-01	60432	Water	pH	Cool 4 deg C	PSEG03	I41	04/01/2025	9040C
Q1683-02	60437	Water	pH	Cool 4 deg C	PSEG03	I41	04/01/2025	9040C
Q1683-03	60438	Water	pH	Cool 4 deg C	PSEG03	I41	04/01/2025	9040C
Q1700-01	32725	Water	pH	Cool 4 deg C	PSEG03	I31	04/02/2025	9040C
Q1700-02	32525	Water	pH	Cool 4 deg C	PSEG03	I31	04/02/2025	9040C
Q1702-03	60442	Water	pH	Cool 4 deg C	PSEG03	L12	04/02/2025	9040C
Q1710-01	34848	Water	pH	Cool 4 deg C	PSEG03	L13	04/02/2025	9040C

Date/Time 04/03/25 08:10

Raw Sample Received by: JB Cope

Raw Sample Relinquished by: JDCSM

Date/Time 04/03/25

Raw Sample Received by: JDCSM

Raw Sample Relinquished by: JB Cope