

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135298

Slope : 99.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/04/2025	08:10
2	CAL2	1	Water	NA	NA	20.2	7.02	04/04/2025	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	04/04/2025	08:12
4	ICV	1	Water	NA	NA	20.2	7.00	04/04/2025	08:15
5	CCV1	1	Water	NA	NA	20.2	2.01	04/04/2025	08:20
6	Q1723-06	1	Water	NA	NA	23.6	6.86	04/04/2025	08:30
7	Q1723-07	1	Water	NA	NA	23.9	6.65	04/04/2025	08:32
8	Q1724-01	1	Water	NA	NA	23.4	5.80	04/04/2025	08:40
9	Q1724-02	1	Water	NA	NA	23.2	5.80	04/04/2025	08:45
10	Q1724-03	1	Water	NA	NA	23.1	5.59	04/04/2025	08:50
11	Q1724-04	1	Water	NA	NA	23.4	5.56	04/04/2025	09:00
12	Q1724-04DUP	1	Water	NA	NA	23.5	5.57	04/04/2025	09:01
13	CCV2	1	Water	NA	NA	20.3	12.02	04/04/2025	09:05

WORKLIST(Hardcopy Internal Chain)

135298

WorkList Name : ph w q1723

WorkList ID : 188718

Department : Wet-Chemistry

Date : 04-04-2025 07:43:40

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1723-06	AUD-0035	Water	pH	Cool 4 deg C	PSEG03	L23	04/03/2025	9040C
Q1723-07	AUD-0036	Water	pH	Cool 4 deg C	PSEG03	L23	04/03/2025	9040C
Q1724-01	BUR-25-0021	Water	pH	Cool 4 deg C	PSEG03	L23	04/03/2025	9040C
Q1724-02	BUR-25-0022	Water	pH	Cool 4 deg C	PSEG03	L23	04/03/2025	9040C
Q1724-03	BUR-25-0023	Water	pH	Cool 4 deg C	PSEG03	L23	04/03/2025	9040C
Q1724-04	BUR-25-0024	Water	pH	Cool 4 deg C	PSEG03	L23	04/03/2025	9040C

Date/Time 04/04/25 08:00

Raw Sample Received by: SB aoc

Raw Sample Relinquished by: SB aoc

Date/Time 04/04/25 12:00

Raw Sample Received by: SB aoc

Raw Sample Relinquished by: SB aoc