

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135300

Slope : 99.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/04/2025	08:10
2	CAL2	1	Water	NA	NA	20.2	7.02	04/04/2025	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	04/04/2025	08:12
4	ICV	1	Water	NA	NA	20.2	7.00	04/04/2025	08:15
5	CCV1	1	Water	NA	NA	20.1	2.02	04/04/2025	09:15
6	Q1721-02	1	Solid	20.02	20	25.3	9.88	04/04/2025	09:25
7	Q1721-02DUP	1	Solid	20.03	20	25.4	9.89	04/04/2025	09:26
8	Q1723-03	1	Solid	20.02	20	25.8	7.91	04/04/2025	09:35
9	Q1723-04	1	Solid	20.01	20	25.4	6.46	04/04/2025	09:44
10	CCV2	1	Water	NA	NA	20.2	12.02	04/04/2025	09:47

WORKLIST(Hardcopy Internal Chain)

135300

WorkList Name : ph s q1723

WorkList ID : 188719

Department : Wet-Chemistry

Date : 04-04-2025 07:44:42

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1721-02	C AU-05-040325	Solid	pH	Cool 4 deg C	PSEG05	L31	04/03/2025	9045D
Q1723-03	A GAS-AUD-25-0041	Solid	pH	Cool 4 deg C	PSEG03	L23	04/03/2025	9045D
Q1723-04	A GAS-AUD-25-0037	Solid	pH	Cool 4 deg C	PSEG03	L23	04/03/2025	9045D

Date/Time 04/04/25 08:00

Raw Sample Received by: JH GOC

Raw Sample Relinquished by: JH GOC

Date/Time 04/04/25 12:00

Raw Sample Received by: JH GOC

Raw Sample Relinquished by: JH GOC