

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB135311

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	04/04/2025	12:00
2	Q1720-01		1	11	100.00	>212.0	>212.0	04/04/2025	12:30
3	Q1720-01DUP		1	11	100.00	>212.0	>212.0	04/04/2025	13:00
4	Q1723-01		1	11	100.00	>212.0	>212.0	04/04/2025	13:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6135311

WorkList Name : fp-4-4+

WorkList ID : 188733

Department : Wet-Chemistry

Date : 04-04-2025 08:31:45

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1720-01	GAS-CAM-03	Solid	Flash Point	Cool 4 deg C	PSEG03	L11	04/03/2025	1010B
Q1723-01	GAS-AUD-25-0040	Solid	Flash Point	Cool 4 deg C	PSEG03	L23	04/03/2025	1010B

Date/Time 04/04/2025 11:50
 Raw Sample Received by: RTH (w)
 Raw Sample Relinquished by: RTH (w) DC

Date/Time 04/04/2025 14:10
 Raw Sample Received by: TH WLO
 Raw Sample Relinquished by: RTH (w)