

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135330

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/07/2025	16:25
2	CAL2	1	Water	NA	NA	20.2	7.01	04/07/2025	16:26
3	CAL3	1	Water	NA	NA	20.2	10.02	04/07/2025	16:30
4	ICV	1	Water	NA	NA	20.3	7.00	04/07/2025	16:33
5	CCV1	1	Water	NA	NA	20.2	2.01	04/07/2025	16:35
6	Q1735-02	1	Water	NA	NA	20.6	7.19	04/07/2025	16:38
7	Q1735-02DUP	1	Water	NA	NA	20.7	7.20	04/07/2025	16:39
8	Q1735-03	1	Water	NA	NA	20.7	7.18	04/07/2025	16:40
9	Q1739-01	1	Water	NA	NA	22.6	6.53	04/07/2025	16:55
10	Q1739-02	1	Water	NA	NA	22.6	6.53	04/07/2025	16:56
11	CCV2	1	Water	NA	NA	20.3	12.02	04/07/2025	17:00

WORKLIST(Hardcopy Internal Chain)

135330

WorkList Name : ph w q1739

WorkList ID : 188786

Department : Wet-Chemistry

Date : 04-07-2025 15:46:10

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1735-02	0401-A	Water	pH	Cool 4 deg C	PSEG03	L31	04/04/2025	9040C
Q1735-03	0401-B	Water	pH	Cool 4 deg C	PSEG03	L31	04/04/2025	9040C
Q1739-01	WC-LIQUID-20250404	Water	pH	Cool 4 deg C	PARS02	L31	04/04/2025	9040C
Q1739-02	WC-LIQUID-20250404	Water	pH	Cool 4 deg C	PARS02	L31	04/04/2025	9040C

Date/Time 04/07/25 16:00
 Raw Sample Received by: gg woc
 Raw Sample Relinquished by: T-C (sw)

Date/Time 04/07/25 18:00
 Raw Sample Received by: T-C (sw)
 Raw Sample Relinquished by: gg woc