

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB135340

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	04/08/2025	08:45
2	Q1735-04		1	14	100.00	>212.0	>212.0	04/08/2025	09:15
3	Q1735-04DUP		1	14	100.00	>212.0	>212.0	04/08/2025	09:45
4	Q1739-02		1	15	100.00	>212.0	>212.0	04/08/2025	10:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB135340

WorkList Name : fp-4-07

WorkList ID : 188757

Department : Wet-Chemistry

Date : 04-07-2025 08:12:38

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1735-04 A	0401-A-0401-B-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	L31	04/04/2025	1010B
Q1739-02 H	WC-LIQUID-20250404	Water	Flash Point	Cool 4 deg C	PARS02	L31	04/04/2025	1010B

Date/Time 04/08/2025 08:35
Raw Sample Received by: RM (wvc)
Raw Sample Relinquished by: RM (wvc)

Date/Time 04/08/2025 10:50
Raw Sample Received by: RM (wvc)
Raw Sample Relinquished by: RM (wvc)