

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135351

Slope : 99.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/09/2025	08:11
2	CAL2	1	Water	NA	NA	20.1	7.01	04/09/2025	08:12
3	CAL3	1	Water	NA	NA	20.2	10.02	04/09/2025	08:15
4	ICV	1	Water	NA	NA	20.2	7.01	04/09/2025	08:20
5	CCV1	1	Water	NA	NA	20.2	2.01	04/09/2025	08:22
6	Q1666-01	1	Water	NA	NA	20.1	6.40	04/09/2025	08:25
7	Q1751-01	1	Water	NA	NA	20.4	7.24	04/09/2025	08:33
8	Q1751-02	1	Water	NA	NA	20.6	7.28	04/09/2025	08:37
9	Q1751-04	1	Water	NA	NA	20.6	7.44	04/09/2025	08:44
10	Q1751-05	1	Water	NA	NA	20.4	7.28	04/09/2025	08:47
11	Q1751-06	1	Water	NA	NA	20.5	7.29	04/09/2025	08:55
12	Q1751-06DUP	1	Water	NA	NA	20.6	7.30	04/09/2025	08:56
13	CCV2	1	Water	NA	NA	20.3	12.02	04/09/2025	09:00

WORKLIST(Hardcopy Internal Chain)

VP (3535)

WorkList Name : ph w q1666 WorkList ID : 188813 Department : Wet-Chemistry Date : 04-09-2025 07:45:47

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1666-01	TW-WTS-05	Water	pH	Cool 4 deg C	ENTA05	I31	03/25/2025	9040C
Q1751-01	0326	Water	pH	Cool 4 deg C	PSEG03	L31	04/08/2025	9040C
Q1751-02	0328	Water	pH	Cool 4 deg C	PSEG03	L31	04/08/2025	9040C
Q1751-04	0407-A	Water	pH	Cool 4 deg C	PSEG03	L31	04/08/2025	9040C
Q1751-05	0407-B	Water	pH	Cool 4 deg C	PSEG03	L31	04/08/2025	9040C
Q1751-06	0408-C	Water	pH	Cool 4 deg C	PSEG03	L31	04/08/2025	9040C

Date/Time 04/09/25 08:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 04/09/25 12:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]