

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB135359

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	04/09/2025	10:20
2	Q1751-01		1	13	100.00	>212.0	>212.0	04/09/2025	10:50
3	Q1751-01DUP		1	13	100.00	>212.0	>212.0	04/09/2025	11:20
4	Q1751-02		1	15	100.00	>212.0	>212.0	04/09/2025	11:50
5	Q1751-03		1	15	100.00	>212.0	>212.0	04/09/2025	12:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6135359

WorkList Name : fp-4-9

WorkList ID : 188819

Department : Wet-Chemistry

Date : 04-09-2025 08:08:32

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1751-01 A	0326	Water	Flash Point	Cool 4 deg C	PSEG03	L31	04/08/2025	1010B
Q1751-02 C	0328	Water	Flash Point	Cool 4 deg C	PSEG03	L31	04/08/2025	1010B
Q1751-03 A	0407-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	L31	04/08/2025	1010B

Date/Time 04/09/2025 10:10
 Raw Sample Received by: RM CWL
 Raw Sample Relinquished by: JRC (CWL)

Date/Time 04/09/2025 13:00
 Raw Sample Received by: JRC (CWL)
 Raw Sample Relinquished by: RM CWL