

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB135362

Slope : 99.2

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/09/2025	15:25
2	CAL2	1	Water	NA	NA	20.3	7.00	04/09/2025	15:26
3	CAL3	1	Water	NA	NA	20.2	10.02	04/09/2025	15:30
4	ICV	1	Water	NA	NA	20.3	7.02	04/09/2025	15:33
5	CCV1	1	Water	NA	NA	20.2	2.01	04/09/2025	15:37
6	Q1756-04	1	Solid	20.02	20	20.9	7.04	04/09/2025	15:50
7	Q1756-04DUP	1	Solid	20.03	20	20.8	7.05	04/09/2025	15:51
8	Q1760-04	1	Solid	20.02	20	23.0	7.55	04/09/2025	16:00
9	Q1760-08	1	Solid	20.02	20	22.5	7.49	04/09/2025	16:10
10	CCV2	1	Water	NA	NA	20.2	12.02	04/09/2025	16:15

WORKLIST(Hardcopy Internal Chain)

135362

WorkList Name : corrosivity p1756

WorkList ID : 188831

Department : Wet-Chemistry

Date : 04-09-2025 14:40:48

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1756-04	TP-9	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	04/09/2025	9045D
Q1760-04	TP-17	Solid	Corrosivity	Cool 4 deg C	PSEG03	L31	04/09/2025	9045D
Q1760-08	TP-15	Solid	Corrosivity	Cool 4 deg C	PSEG03	L31	04/09/2025	9045D

Date/Time 04/09/25 15:10

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 04/09/25

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]