

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135386

Slope : 98.5

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/11/2025	08:14
2	CAL2	1	Water	NA	NA	20.3	7.00	04/11/2025	08:15
3	CAL3	1	Water	NA	NA	20.2	10.02	04/11/2025	08:20
4	ICV	1	Water	NA	NA	20.2	7.00	04/11/2025	08:22
5	CCV1	1	Water	NA	NA	20.2	2.01	04/11/2025	08:25
6	Q1764-01	1	Solid	20.02	20	24.6	10.02	04/11/2025	08:35
7	Q1764-01DUP	1	Solid	20.03	20	24.7	10.03	04/11/2025	08:36
8	CCV2	1	Water	NA	NA	20.3	12.02	04/11/2025	08:40

WORKLIST(Hardcopy Internal Chain)

WB 135386

WorkList Name : ph q1764 s

WorkList ID : 188857

Department : Wet-Chemistry

Date : 04-11-2025 07:51:27

Sample

Customer Sample

Matrix

Test

Preservative

Customer

Raw Sample
Storage
Location

Collect Date

Method

Q1764-01 BUR-25-0020

Solid pH

Cool 4 deg C

PSEG03

L41

04/10/2025

9045D

Date/Time 04-11-25 08:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 04-11-25 12:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]