

LB135395


7471B

INSTRUMENT ID: (V1

Linear

μ Abs.:

50300



A= 0.0000e+000

B= 2.0104e-004

C= -5.3578e-004

Rho= 0.9996644

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.013	0.013	66	0.000	66					0/27
0.2	0.200	0.220	0.020	1098	0.0 %	1098					10
2.5	2.500	2.487	-0.013	12372	0.0 %	12372					-1
5.0	5.000	5.063	0.063	25187	0.0 %	25187					-1
7.5	7.500	7.306	-0.194	36344	0.0 %	36344					-3
10.0	10.000	10.112	0.112	50300	0.0 %	50300					1

LB135395 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 11 Apr 2025 12:29:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	66	-	0.000	7471B	PPB	11 Apr 2025 12:31:11	S	Std
0.2 - 1	50-2	1098	-	0.200	7471B	PPB	11 Apr 2025 12:33:29	S	Std
2.5 - 1	52.5	12372	-	2.500	7471B	PPB	11 Apr 2025 12:35:47	S	Std
5.0 - 1	55	25187	-	5.000	7471B	PPB	11 Apr 2025 12:38:03	S	Std
7.5 - 1	57.5	36344	-	7.500	7471B	PPB	11 Apr 2025 12:40:22	S	Std
10.0 - 1	510	50300	-	10.000	7471B	PPB	11 Apr 2025 12:42:42	S	Std
ICV85 - 1	ICV85	18268	3.6720	-	7471B	PPB	11 Apr 2025 12:45:42	U	SMPL
ICB85 - 1	ICB85	-260	-0.0528	-	7471B	PPB	11 Apr 2025 12:47:58	U	SMPL
CCV63 - 1	CCV63	25074	5.0402	-	7471B	PPB	11 Apr 2025 12:50:16	U	SMPL
CCB63 - 1	CCB63	-382	-0.0773	-	7471B	PPB	11 Apr 2025 12:52:32	U	SMPL
CRA - 1	CRA	1120	0.2246	-	7471B	PPB	11 Apr 2025 12:54:50	U	SMPL
HighStd - 1	HighStd	50604	10.1727	-	7471B	PPB	11 Apr 2025 12:57:08	U	SMPL
ChkStd - 1	ChkStd	31794	6.3912	-	7471B	PPB	11 Apr 2025 12:59:21	U	SMPL
PB167572BL - 1	PBS	-452	-0.0914	-	7471B	PPB	11 Apr 2025 13:01:46	U	SMPL
Q1754-01 - 1	TP-1	305942	61.5047	-	7471B	PPB	11 Apr 2025 13:06:25	U	SMPL
Q1754-03 - 1	TP-1-CONCRETE	26136	5.2537	-	7471B	PPB	11 Apr 2025 13:08:44	U	SMPL
Q1756-01 - 1	TP-9	4710	0.9463	-	7471B	PPB	11 Apr 2025 13:11:58	U	SMPL
Q1760-01 - 1	TP-17	767	0.1537	-	7471B	PPB	11 Apr 2025 13:14:24	U	SMPL
Q1760-05 - 1	TP-15	1379	0.2767	-	7471B	PPB	11 Apr 2025 13:16:41	U	SMPL
CCV64 - 1	CCV64	24993	5.0239	-	7471B	PPB	11 Apr 2025 13:18:57	U	SMPL
CCB64 - 1	CCB64	-449	-0.0908	-	7471B	PPB	11 Apr 2025 13:21:12	U	SMPL
Q1768-01 - 1	CRUSHED STONE	2024	0.4064	-	7471B	PPB	11 Apr 2025 13:23:31	U	SMPL
Q1771-01 - 1	TP-6C	11329	2.2770	-	7471B	PPB	11 Apr 2025 13:25:46	U	SMPL
Q1771-05 - 1	TP-7C	2918	0.5861	-	7471B	PPB	11 Apr 2025 13:28:04	U	SMPL
Q1771-09 - 1	TP-3A	660	0.1321	-	7471B	PPB	11 Apr 2025 13:30:22	U	SMPL
Q1779-01 - 1	TP-22	774	0.1551	-	7471B	PPB	11 Apr 2025 13:32:40	U	SMPL
Q1779-05 - 1	TP-23	3127	0.6281	-	7471B	PPB	11 Apr 2025 13:34:56	U	SMPL
Q1781-01 - 1	WC-13	1787	0.3587	-	7471B	PPB	11 Apr 2025 13:37:12	U	SMPL
Q1781-05 - 1	WC-12	242	0.0481	-	7471B	PPB	11 Apr 2025 13:39:30	U	SMPL
Q1781-09 - 1	WC-11	843	0.1689	-	7471B	PPB	11 Apr 2025 13:41:47	U	SMPL
Q1781-09DUP - 1	WC-11DUP	1009	0.2023	-	7471B	PPB	11 Apr 2025 13:44:03	U	SMPL
CCV65 - 1	CCV65	24926	5.0105	-	7471B	PPB	11 Apr 2025 13:46:18	U	SMPL
CCB65 - 1	CCB65	-399	-0.0807	-	7471B	PPB	11 Apr 2025 13:48:34	U	SMPL
Q1781-09MS - 1	WC-11MS	17393	3.4961	-	7471B	PPB	11 Apr 2025 13:50:53	U	SMPL
Q1781-09MSD - 1	WC-11MSD	17738	3.5654	-	7471B	PPB	11 Apr 2025 13:55:58	U	SMPL
Q1781-09LX5 - 1		-34	-0.0074	-	7471B	PPB	11 Apr 2025 13:58:13	U	SMPL
Q1781-09A - 1		18893	3.7976	-	7471B	PPB	11 Apr 2025 14:00:31	U	SMPL
Q1754-01DLX10 - 1		26643	5.3557	-	7471B	PPB	11 Apr 2025 14:02:47	U	SMPL
PB167572BNS - 1	LCSS	18238	3.6659	-	7471B	PPB	11 Apr 2025 14:10:16	U	SMPL
CCV66 - 1	CCV66	24838	4.9928	-	7471B	PPB	11 Apr 2025 14:13:00	U	SMPL
CCB66 - 1	CCB66	-390	-0.0789	-	7471B	PPB	11 Apr 2025 14:15:20	U	SMPL