

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB135397

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	04/11/2025	13:20
2	Q1763-02		1	14	100.00	>212.0	>212.0	04/11/2025	13:50
3	Q1763-02DUP		1	14	100.00	>212.0	>212.0	04/11/2025	14:20
4	Q1764-02		1	13	100.00	>212.0	>212.0	04/11/2025	14:50
5	Q1767-01		1	16	100.00	>212.0	>212.0	04/11/2025	15:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6135397

WorkList Name : fp-04-10 WorkList ID : 188836 Department : Wet-Chemistry Date : 04-10-2025 08:15:02

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1763-02 C	AUD-25-0047	Water	Flash Point	Cool 4 deg C	PSEG03	I31	04/10/2025	1010B
Q1764-02 C	BUR-25-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	L41	04/10/2025	1010B
Q1767-01 E	NWB-2152	Water	Flash Point	Cool 4 deg C	PSEG03	F11	04/10/2025	1010B

Date/Time 04/11/2025 13:05
Raw Sample Received by: RM CWJ
Raw Sample Relinquished by: JH GEDC

Date/Time 04/11/2025 16:00
Raw Sample Received by: JH GEDC
Raw Sample Relinquished by: RM CWJ