

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB135400

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/11/2025	14:00
2	CAL2	1	Water	NA	NA	20.2	7.00	04/11/2025	14:01
3	CAL3	1	Water	NA	NA	20.1	10.02	04/11/2025	14:05
4	ICV	1	Water	NA	NA	20.2	7.01	04/11/2025	14:10
5	CCV1	1	Water	NA	NA	20.3	2.01	04/11/2025	14:15
6	Q1787-12	1	Solid	20.02	20	23.2	10.15	04/11/2025	14:30
7	Q1788-04	1	Solid	20.03	20	22.7	9.42	04/11/2025	14:40
8	Q1791-07	1	Solid	20.02	20	22.6	10.65	04/11/2025	14:55
9	Q1791-07DUP	1	Solid	20.03	20	22.7	10.66	04/11/2025	14:56
10	CCV2	1	Water	NA	NA	20.3	12.02	04/11/2025	15:00

WORKLIST(Hardcopy Internal Chain)

WB 135400

WorkList Name : corrosivity Q1768 WorkList ID : 188884 Department : Wet-Chemistry Date : 04-11-2025 11:55:52

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1787-04	TP-5C	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	04/10/2025	9045D
Q1787-08	TP-2C	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	04/10/2025	9045D
Q1787-12	TP-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	04/10/2025	9045D
Q1788-04	WC-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	L31	04/11/2025	9045D
Q1791-07	290502	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	04/11/2025	9045D

Date/Time 04-11-25 13:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 04-11-25 17:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]