

LB 135413

7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

49973

A= 0.0000e+000

B= 1.9803e-004

C= -4.0457e-002

Rho= 0.9996835

Accept=Accepted

Max. Conc: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	4/15
0.0	0.000	-0.052	-0.052	-60	0.000	-60					-
0.2	0.200	0.171	-0.029	1068	0.0 %	1068					-15
2.5	2.500	2.494	-0.006	12798	0.0 %	12798					0
5.0	5.000	5.130	0.130	26110	0.0 %	26110					3
7.5	7.500	7.602	0.102	38591	0.0 %	38591					1
10.0	10.000	9.856	-0.144	49973	0.0 %	49973					-1

LB135413 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 Apr 2025 13:50:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30	-60	-	0.00007471B	PPB		14 Apr 2025 13:50:50	S	Std
0.2 - 1	50.2	1068	-	0.20007471B	PPB		14 Apr 2025 13:53:07	S	Std
2.5 - 1	52.5	12798	-	2.50007471B	PPB		14 Apr 2025 13:55:25	S	Std
5.0 - 1	55	26110	-	5.00007471B	PPB		14 Apr 2025 13:57:42	S	Std
7.5 - 1	57.5	38591	-	7.50007471B	PPB		14 Apr 2025 14:00:01	S	Std
10.0 - 1	50	49973	-	10.00007471B	PPB		14 Apr 2025 14:02:20	S	Std
ICV86 - 1	ICV86	19683	3.8573	-7471B	PPB		14 Apr 2025 14:07:57	U	SMPL
ICB86 - 1	ICB86	-62	-0.0527	-7471B	PPB		14 Apr 2025 14:12:51	U	SMPL
CCV67 - 1	CCV67	26460	5.1994	-7471B	PPB		14 Apr 2025 14:15:06	U	SMPL
CCB67 - 1	CCB67	-290	-0.0979	-7471B	PPB		14 Apr 2025 14:17:21	U	SMPL
CRA - 1	CRA	1078	0.1730	-7471B	PPB		14 Apr 2025 14:19:40	U	SMPL
HighStd - 1	HighStd	50982	10.0554	-7471B	PPB		14 Apr 2025 14:21:56	U	SMPL
ChkStd - 1	ChkStd	36120	7.1123	-7471B	PPB		14 Apr 2025 14:24:12	U	SMPL
PB167588BL - 1	PBS	-98	-0.0599	-7471B	PPB		14 Apr 2025 14:26:38	U	SMPL
PB167588BS - 1	LCSS	22300	4.3756	-7471B	PPB		14 Apr 2025 14:29:01	U	SMPL
Q1783-01 - 1	SU-03-41125	3495	0.6517	-7471B	PPB		14 Apr 2025 14:31:18	U	SMPL
Q1784-01 - 1	OR-01-41125	2917	0.5372	-7471B	PPB		14 Apr 2025 14:33:38	U	SMPL
Q1785-01 - 1	HR-02-041125	3255	0.6041	-7471B	PPB		14 Apr 2025 14:35:57	U	SMPL
Q1785-03 - 1	HR-03-041125	2642	0.4827	-7471B	PPB		14 Apr 2025 14:38:16	U	SMPL
Q1787-09 - 1	TP-1	11550	2.2468	-7471B	PPB		14 Apr 2025 14:40:33	U	SMPL
CCV68 - 1	CCV68	26061	5.1204	-7471B	PPB		14 Apr 2025 14:42:48	U	SMPL
CCB68 - 1	CCB68	-291	-0.0981	-7471B	PPB		14 Apr 2025 14:45:06	U	SMPL
Q1788-01 - 1	WC-1	24515	4.8142	-7471B	PPB		14 Apr 2025 14:47:27	U	SMPL
Q1791-07 - 1	290502	242	0.0075	-7471B	PPB		14 Apr 2025 14:49:42	U	SMPL
Q1792-01 - 1	TR-05-041125	5204	0.9901	-7471B	PPB		14 Apr 2025 14:56:53	U	SMPL
Q1792-01DUP - 1	TR-05-041125DUP	4982	0.9461	-7471B	PPB		14 Apr 2025 14:59:08	U	SMPL
Q1792-01MS - 1	TR-05-041125MS	25887	5.0859	-7471B	PPB		14 Apr 2025 15:01:25	U	SMPL
Q1792-01MSD - 1	TR-05-041125MSD	24797	4.8701	-7471B	PPB		14 Apr 2025 15:03:44	U	SMPL
Q1792-01LX5 - 1		817	0.1213	-7471B	PPB		14 Apr 2025 15:06:03	U	SMPL
Q1792-01A - 1		26058	5.1198	-7471B	PPB		14 Apr 2025 15:08:23	U	SMPL
CCV69 - 1	CCV69	26087	5.1255	-7471B	PPB		14 Apr 2025 15:10:40	U	SMPL
CCB69 - 1	CCB69	-209	-0.0818	-7471B	PPB		14 Apr 2025 15:13:00	U	SMPL