

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB135414

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	04/14/2025	12:45
2	Q1795-02		1	15	100.00	>212.0	>212.0	04/14/2025	13:15
3	Q1795-02DUP		1	15	100.00	>212.0	>212.0	04/14/2025	13:45
4	Q1795-03		1	13	100.00	>212.0	>212.0	04/14/2025	14:15
5	Q1798-01		1	12	100.00	>212.0	>212.0	04/14/2025	14:45
6	Q1799-01		1	15	100.00	>212.0	>212.0	04/14/2025	15:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6135414

WorkList Name : fp-04-14 WorkList ID : 188898 Department : Wet-Chemistry Date : 04-14-2025 08:54:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1795-02	31125	Water	Flash Point	Cool 4 deg C	PSEG03	L41	04/11/2025	1010B
Q1795-03	10824-A	Water	Flash Point	Cool 4 deg C	PSEG03	L41	04/11/2025	1010B
Q1798-01	TOTE(COMP)-1799	Water	Flash Point	Cool 4 deg C	PSEG03	L41	04/11/2025	1010B
Q1799-01	FRAC-TANK-216-PPR	Water	Flash Point	Cool 4 deg C	PSEG03	L41	04/11/2025	1010B

Date/Time 04/14/2025 12:35
 Raw Sample Received by: RT we
 Raw Sample Relinquished by: RT we

Date/Time 04/14/2025 16:00
 Raw Sample Received by: RT we
 Raw Sample Relinquished by: RT we