

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135421

Slope : 99.2

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/14/2025	15:30
2	CAL2	1	Water	NA	NA	20.2	7.00	04/14/2025	15:31
3	CAL3	1	Water	NA	NA	20.3	10.02	04/14/2025	15:33
4	ICV	1	Water	NA	NA	20.3	7.02	04/14/2025	15:35
5	CCV1	1	Water	NA	NA	20.2	2.01	04/14/2025	15:40
6	Q1800-02	1	Solid	20.02	20	24.7	12.25	04/15/2025	16:00
7	Q1800-02DUP	1	Solid	20.03	20	24.8	12.26	04/14/2025	16:01
8	CCV2	1	Water	NA	NA	20.3	12.02	04/14/2025	16:30

WORKLIST(Hardcopy Internal Chain)

UB 135421

WorkList Name : ph q1800 s WorkList ID : 188910 Department : Wet-Chemistry Date : 04-14-2025 13:51:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1800-02	WC-A4-01-C	Solid	pH	Cool 4 deg C	ENTA05	L41	04/11/2025	9045D

Date/Time 04-14-25 15:20
Raw Sample Received by: SA WOC
Raw Sample Relinquished by: SA WOC

Date/Time 04-14-25 18:00
Raw Sample Received by: SA WOC
Raw Sample Relinquished by: SA WOC