

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135428

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/15/2025	08:15
2	CAL2	1	Water	NA	NA	20.2	7.00	04/15/2025	08:16
3	CAL3	1	Water	NA	NA	20.2	10.02	04/15/2025	08:20
4	ICV	1	Water	NA	NA	20.2	7.00	04/15/2025	08:22
5	CCV1	1	Water	NA	NA	20.3	2.01	04/15/2025	08:25
6	Q1810-01	1	Water	NA	NA	20.3	7.69	04/15/2025	08:35
7	Q1810-01DUP	1	Water	NA	NA	20.4	7.70	04/15/2025	08:36
8	Q1810-02	1	Water	NA	NA	20.7	9.55	04/15/2025	08:45
9	CCV2	1	Water	NA	NA	20.2	12.02	04/15/2025	08:50

WORKLIST(Hardcopy Internal Chain)

WB 135428

WorkList Name : PH Q1810 W

WorkList ID : 188918

Department : Wet-Chemistry

Date : 04-15-2025 08:05:19

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1810-01	P MOO-25-0117	Water	pH	Cool 4 deg C	PSEG03	L51	04/14/2025	9040C
Q1810-02	M MOO-25-0118	Water	pH	Cool 4 deg C	PSEG03	L51	04/14/2025	9040C

Date/Time 04-15-25 08:10

Raw Sample Received by: JP WOI

Raw Sample Relinquished by: JP SW

Date/Time 04-15-25

Raw Sample Received by: JP SW

Raw Sample Relinquished by: JP WOI