

LB135432


245.1

INSTRUMENT ID: CV4

Linear

 μ Abs.:

42241



A= 0.0000e+000

B= 2.3493e-004

C= -3.1518e-003

Rho= 0.9999226

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/1
0.0	0.000	-0.016	-0.016	-55	0.000	-55					1
0.05	0.050	0.049	-0.001	220	0.0 %	220	0				-2
0.2	0.200	0.173	-0.027	749	0.0 %	749					-14
2.5	2.500	2.501	0.001	10658	0.0 %	10658					0
5.0	5.000	5.057	0.057	21539	0.0 %	21539					1
7.5	7.500	7.566	0.066	32219	0.0 %	32219					1
10.0	10.000	9.921	-0.079	42241	0.0 %	42241					1

LB135432 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 15 Apr 2025 09:15:48

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-55	-	0.0000	245.1	PPB	15 Apr 2025 09:32:44	S	Std
0.05 - 1	50-45	220	-	0.0500	245.1	PPB	15 Apr 2025 09:37:54	S	Std
0.2 - 1	50-2	749	-	0.2000	245.1	PPB	15 Apr 2025 09:40:11	S	Std
2.5 - 1	52-15	10658	-	2.5000	245.1	PPB	15 Apr 2025 09:42:28	S	Std
5.0 - 1	55	21539	-	5.0000	245.1	PPB	15 Apr 2025 09:44:44	S	Std
7.5 - 1	52.5	32219	-	7.5000	245.1	PPB	15 Apr 2025 09:50:53	S	Std
10.0 - 1	510	42241	-	10.0000	245.1	PPB	15 Apr 2025 09:58:39	S	Std
ICV88 - 1	ICV88	16890	3.9649	-245.1		PPB	15 Apr 2025 10:03:10	U	SMPL
ICB88 - 1	ICB88	-330	-0.0807	-245.1		PPB	15 Apr 2025 10:05:24	U	SMPL
CCV72 - 1	CCV72	21603	5.0721	-245.1		PPB	15 Apr 2025 10:07:42	U	SMPL
CCB72 - 1	CCB72	-393	-0.0955	-245.1		PPB	15 Apr 2025 10:09:57	U	SMPL
CRA - 1	CRA	278	0.0622	-245.1		PPB	15 Apr 2025 10:20:07	U	SMPL
HighStd - 1	HighStd	42685	10.0250	-245.1		PPB	15 Apr 2025 10:22:22	U	SMPL
ChkStd - 1	ChkStd	28771	6.7561	-245.1		PPB	15 Apr 2025 10:24:38	U	SMPL
PB167601BL - 1	PBW	-175	-0.0443	-245.1		PPB	15 Apr 2025 10:26:57	U	SMPL
PB167601BS - 1	LCSW	15766	3.7008	-245.1		PPB	15 Apr 2025 10:34:52	U	SMPL
Q1703-02 - 1	COMP	716	0.1651	-245.1		PPB	15 Apr 2025 10:37:08	U	SMPL
Q1703-02DUP - 1	COMPDUP	776	0.1792	-245.1		PPB	15 Apr 2025 10:39:27	U	SMPL
Q1703-02MS - 1	COMPMS	15378	3.6097	-245.1		PPB	15 Apr 2025 10:41:45	U	SMPL
Q1703-02MSD - 1	COMPMSD	14616	3.4306	-245.1		PPB	15 Apr 2025 10:46:48	U	SMPL
CCV73 - 1	CCV73	20923	4.9124	-245.1		PPB	15 Apr 2025 10:49:03	U	SMPL
CCB73 - 1	CCB73	-302	-0.0741	-245.1		PPB	15 Apr 2025 10:51:20	U	SMPL
Q1703-02LX5 - 1		68	0.0128	-245.1		PPB	15 Apr 2025 10:53:37	U	SMPL
Q1703-02A - 1		19365	4.5463	-245.1		PPB	15 Apr 2025 10:55:52	U	SMPL
CCV74 - 1	CCV74	20868	4.8994	-245.1		PPB	15 Apr 2025 10:58:07	U	SMPL
CCB74 - 1	CCB74	-330	-0.0807	-245.1		PPB	15 Apr 2025 11:00:24	U	SMPL