

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB135436

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	83.0	04/15/2025	10:15
2	Q1791-09		1	13	100.00	>212.0	>212.0	04/15/2025	10:45
3	Q1791-09DUP		1	12	100.00	>212.0	>212.0	04/15/2025	11:15
4	Q1791-10		1	13	100.00	>212.0	>212.0	04/15/2025	11:45
5	Q1791-11		1	14	100.00	>212.0	>212.0	04/15/2025	12:15
6	Q1791-12		1	12	100.00	>212.0	>212.0	04/15/2025	12:45
7	Q1791-13		1	12	100.00	>212.0	>212.0	04/15/2025	13:15
8	Q1791-14		1	13	100.00	>212.0	>212.0	04/15/2025	13:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB135436

WorkList Name : fp-4-15

WorkList ID : 188924

Department : Wet-Chemistry

Date : 04-15-2025 08:18:04

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1791-09	VB16193	Water	Flash Point	Cool 4 deg C	PSEG03	L31	04/11/2025	1010B
Q1791-10	280722	Water	Flash Point	Cool 4 deg C	PSEG03	L31	04/11/2025	1010B
Q1791-11	280791	Water	Flash Point	Cool 4 deg C	PSEG03	L31	04/11/2025	1010B
Q1791-12	V1207	Water	Flash Point	Cool 4 deg C	PSEG03	L31	04/11/2025	1010B
Q1791-13	279312	Water	Flash Point	Cool 4 deg C	PSEG03	L31	04/11/2025	1010B
Q1791-14	VB16164	Water	Flash Point	Cool 4 deg C	PSEG03	L31	04/11/2025	1010B

Date/Time 04/15/25 10:05
 Raw Sample Received by: 12100
 Raw Sample Relinquished by: H(000)

Date/Time 04/15/25 14:30
 Raw Sample Received by: H(000)
 Raw Sample Relinquished by: 12100