

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB135441

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	83.0	04/15/2025	09:10
2	Q1794-04		1	12	100.00	>212.0	>212.0	04/15/2025	09:40
3	Q1794-04DUP		1	12	100.00	>212.0	>212.0	04/15/2025	10:10
4	Q1797-01		1	15	100.00	>212.0	>212.0	04/15/2025	10:50
5	Q1810-01		1	13	100.00	>212.0	>212.0	04/15/2025	11:20
6	Q1810-02		1	15	100.00	>212.0	>212.0	04/15/2025	12:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6135441

WorkList Name : flash point-04-25

WorkList ID : 138945

Department : Wet-Chemistry

Date : 04-15-2025 08:42:52

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1794-04 E	40325-A-B-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	L41	04/11/2025	1010B
Q1797-01 D	SEAL-POT-40125	Water	Flash Point	Cool 4 deg C	PSEG03	L41	04/11/2025	1010B
Q1810-01 A	MOO-25-0117	Water	Flash Point	Cool 4 deg C	PSEG03	L51	04/14/2025	1010B
Q1810-02 Q	MOO-25-0118	Water	Flash Point	Cool 4 deg C	PSEG03	L51	04/14/2025	1010B

Date/Time 04/15/2025 09:00
 Raw Sample Received by: RM CWL2
 Raw Sample Relinquished by: JP (CWL)

Date/Time 04/15/2025 13:20
 Raw Sample Received by: JP (CWL)
 Raw Sample Relinquished by: RM CWL2