

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135464

Slope : 99.2

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/17/2025	08:30
2	CAL2	1	Water	NA	NA	20.3	7.00	04/17/2025	08:31
3	CAL3	1	Water	NA	NA	20.3	10.02	04/17/2025	08:35
4	ICV	1	Water	NA	NA	20.3	7.01	04/17/2025	08:40
5	CCV1	1	Water	NA	NA	20.2	2.01	04/17/2025	08:45
6	Q1824-01	1	Solid	20.02	20	24.6	5.93	04/17/2025	09:00
7	Q1824-01DUP	1	Solid	20.03	20	24.8	5.94	04/17/2025	09:01
8	CCV2	1	Water	NA	NA	20.3	12.02	04/17/2025	09:05

WORKLIST(Hardcopy Internal Chain)

B 135464

WorkList Name : PH S Q1814

WorkList ID : 188974

Department : Wet-Chemistry

Date : 04-17-2025 08:16:36

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1824-01 - B	284	Solid	pH	Cool 4 deg C	PSEG03	L31	04/16/2025	9045D

Date/Time 04-17-25 08:25

Raw Sample Received by: JB WOC

Raw Sample Relinquished by: JB WOC

Date/Time 04-17-25 11:30

Raw Sample Received by: JB WOC

Raw Sample Relinquished by: JB WOC