

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135465

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/17/2025	08:30
2	CAL2	1	Water	NA	NA	20.3	7.00	04/17/2025	08:31
3	CAL3	1	Water	NA	NA	20.3	10.02	04/17/2025	08:35
4	ICV	1	Water	NA	NA	20.3	7.01	04/17/2025	08:40
5	CCV1	1	Water	NA	NA	20.2	2.01	04/17/2025	08:45
6	Q1814-01	1	Water	NA	NA	20.6	6.99	04/17/2025	09:20
7	Q1814-01DUP	1	Water	NA	NA	20.7	7.00	04/17/2025	09:21
8	Q1814-03	1	Water	NA	NA	21.3	5.30	04/17/2025	09:30
9	Q1816-01	1	Water	NA	NA	20.7	8.27	04/17/2025	09:33
10	Q1818-05	1	Water	NA	NA	20.7	10.82	04/17/2025	09:47
11	Q1818-06	1	Water	NA	NA	20.7	10.11	04/17/2025	09:55
12	Q1818-07	1	Water	NA	NA	20.9	6.96	04/17/2025	10:10
13	Q1821-03	1	Water	NA	NA	22.6	5.92	04/17/2025	10:15
14	Q1824-04	1	Water	NA	NA	20.9	7.01	04/17/2025	10:20
15	CCV2	1	Water	NA	NA	20.3	12.02	04/17/2025	10:22

WORKLIST(Hardcopy Internal Chain)

135465

WorkList Name : PH W Q1821

WorkList ID : 188973

Department : Wet-Chemistry

Date : 04-17-2025 08:16:04

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1814-01	B A3729	Water	pH	Cool 4 deg C	PSEG03		04/16/2025	9040C
Q1814-03	C SCRUBBLE-TANK-LIQUID	Water	pH	Cool 4 deg C	PSEG03	L41	04/16/2025	9040C
Q1816-01	C 441	Water	pH	Cool 4 deg C	PSEG03	L41	04/16/2025	9040C
Q1818-05	C 3828	Water	pH	Cool 4 deg C	PSEG03	L41	04/16/2025	9040C
Q1818-06	C 3829	Water	pH	Cool 4 deg C	PSEG03	L41	04/16/2025	9040C
Q1818-07	C 3651	Water	pH	Cool 4 deg C	PSEG03	L41	04/16/2025	9040C
Q1821-03	D GW-DRUM	Water	pH	Cool 4 deg C	PSEG03	N31	04/16/2025	9040C
Q1824-04	C 285	Water	pH	Cool 4 deg C	PSEG03	L31	04/16/2025	9040C

Date/Time 04-17-25 08:25

Raw Sample Received by: 80 WOC

Raw Sample Relinquished by: JWSH

Date/Time 04-17-25

Raw Sample Received by: JWSH

Raw Sample Relinquished by: 80 WOC