

LB135470


7470A

INSTRUMENT ID: CV3

Linear

 μ Abs.:

47703



A= 0.0000e+000

B= 2.0848e-004

C= 1.0607e-002

Rho= 0.9992703

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	6700
0.0	0.000	-0.011	-0.011	-102	0.000	-102	0				-15
0.2	0.200	0.170	-0.030	763	0.0 %	763					7
2.5	2.500	2.685	0.185	12828	0.0 %	12828					-5
5.0	5.000	4.754	-0.246	22752	0.0 %	22752					2
7.5	7.500	7.646	0.146	36626	0.0 %	36626					0
10.0	10.000	9.956	-0.044	47703	0.0 %	47703					

LB135470 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Apr 2025 11:43:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-102	-	0.0000	7470A	PPB	17 Apr 2025 11:53:36	S	Std
0.2 - 1	50.2	763	-	0.2000	7470A	PPB	17 Apr 2025 11:55:54	S	Std
2.5 - 1	52.5	12828	-	2.5000	7470A	PPB	17 Apr 2025 11:58:11	S	Std
5.0 - 1	55	22752	-	5.0000	7470A	PPB	17 Apr 2025 12:00:28	S	Std
7.5 - 1	55	36626	-	7.5000	7470A	PPB	17 Apr 2025 12:04:56	S	Std
10.0 - 1	516	47703	-	10.0000	7470A	PPB	17 Apr 2025 12:09:53	S	Std
ICV92 - 1	ICV92	20481	4.2805	-	7470A	PPB	17 Apr 2025 12:15:31	U	SMPL
ICB92 - 1	ICB92	-197	-0.0305	-	7470A	PPB	17 Apr 2025 12:17:46	U	SMPL
CCV85 - 1	CCV85	25251	5.2749	-	7470A	PPB	17 Apr 2025 12:20:03	U	SMPL
CCB85 - 1	CCB85	-294	-0.0507	-	7470A	PPB	17 Apr 2025 12:22:18	U	SMPL
CRA - 1	CRA	730	0.1628	-	7470A	PPB	17 Apr 2025 12:24:35	U	SMPL
HighStd - 1	HighStd	45648	9.5273	-	7470A	PPB	17 Apr 2025 12:26:50	U	SMPL
ChkStd - 1	ChkStd	31963	6.6742	-	7470A	PPB	17 Apr 2025 12:29:06	U	SMPL
PB167634BL - 1	PBW	-414	-0.0757	-	7470A	PPB	17 Apr 2025 12:31:24	U	SMPL
PB167634BS - 1	LCSW	19039	3.9798	-	7470A	PPB	17 Apr 2025 12:36:28	U	SMPL
Q1812-02 - 1	RINSE-EB-TANK-041525	-188	-0.0286	-	7470A	PPB	17 Apr 2025 12:38:44	U	SMPL
Q1812-03 - 1	RINSE-EB-PUMP-041525	172	0.0465	-	7470A	PPB	17 Apr 2025 12:41:02	U	SMPL
Q1814-01 - 1	A3729	4372	0.9221	-	7470A	PPB	17 Apr 2025 12:43:19	U	SMPL
Q1814-02 - 1	SCRUBBLE-TANK-LIQUID	7921	1.6620	-	7470A	PPB	17 Apr 2025 12:45:38	U	SMPL
Q1816-01 - 1	441	-449	-0.0830	-	7470A	PPB	17 Apr 2025 12:47:54	U	SMPL
CCV86 - 1	CCV86	22256	4.6505	-	7470A	PPB	17 Apr 2025 12:53:40	U	SMPL
CCB86 - 1	CCB86	-317	-0.0555	-	7470A	PPB	17 Apr 2025 12:55:57	U	SMPL
Q1818-05 - 1	3828	490	0.1128	-	7470A	PPB	17 Apr 2025 12:58:15	U	SMPL
Q1818-06 - 1	3829	543	0.1238	-	7470A	PPB	17 Apr 2025 13:00:30	U	SMPL
Q1818-07 - 1	3651	-64	-0.0027	-	7470A	PPB	17 Apr 2025 13:02:48	U	SMPL
Q1824-04 - 1	285	273	0.0675	-	7470A	PPB	17 Apr 2025 13:05:04	U	SMPL
Q1824-04DUP - 1	285DUP	188	0.0498	-	7470A	PPB	17 Apr 2025 13:07:20	U	SMPL
Q1824-04MS - 1	285MS	16873	3.5283	-	7470A	PPB	17 Apr 2025 13:14:15	U	SMPL
Q1824-04MSD - 1	285MSD	16206	3.3892	-	7470A	PPB	17 Apr 2025 13:16:31	U	SMPL
Q1824-04LX5 - 1		-193	-0.0296	-	7470A	PPB	17 Apr 2025 13:18:50	U	SMPL
Q1824-04A - 1		18409	3.8485	-	7470A	PPB	17 Apr 2025 13:21:09	U	SMPL
CCV87 - 1	CCV87	23447	4.8988	-	7470A	PPB	17 Apr 2025 13:23:26	U	SMPL
CCB87 - 1	CCB87	-323	-0.0567	-	7470A	PPB	17 Apr 2025 13:25:44	U	SMPL