

LB135-480


7471B

INSTRUMENT 1A/CV1

Linear

 μ Abs.:

51186



A= 0.0000e+000

B= 1.9638e-004

C= -2.5668e-002

Rho= 0.9999739

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.015	0.015	209	0.000	209					
0.2	0.200	0.186	-0.014	1079	0.0 %	1079					
2.5	2.500	2.499	-0.001	12857	0.0 %	12857					
5.0	5.000	5.024	0.024	25712	0.0 %	25712					
7.5	7.500	7.449	-0.051	38063	0.0 %	38063					
10.0	10.000	10.026	0.026	51186	0.0 %	51186					

0.0
1.7
0.0
1.0
0

LB135480 INSTRUMENT ID

Method: 7471B

Operator: Admin

Date of Analysis: 18 Apr 2025 08:56:53

Sample ID	Extended ID	Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	10	209	-	0.0000	7471B	PPB	18 Apr 2025 08:57:09	S	Std
0.2 - 1	50-2	1079	-	0.2000	7471B	PPB	18 Apr 2025 08:59:26	S	Std
2.5 - 1	52.5	12857	-	2.5000	7471B	PPB	18 Apr 2025 09:01:43	S	Std
5.0 - 1	55	25712	-	5.0000	7471B	PPB	18 Apr 2025 09:03:59	S	Std
7.5 - 1	57.5	38063	-	7.5000	7471B	PPB	18 Apr 2025 09:06:19	S	Std
10.0 - 1	110	51186	-	10.0000	7471B	PPB	18 Apr 2025 09:08:37	S	Std
ICV94 - 1	ICV94	20464	3.9931	-	7471B	PPB	18 Apr 2025 09:11:33	U	SMPL
ICB94 - 1	ICB94	65	-0.0129	-	7471B	PPB	18 Apr 2025 09:13:49	U	SMPL
CCV92 - 1	CCV92	25043	4.8923	-	7471B	PPB	18 Apr 2025 09:16:07	U	SMPL
CCB92 - 1	CCB92	24	-0.0210	-	7471B	PPB	18 Apr 2025 09:18:22	U	SMPL
CRA - 1	CRA	950	0.1609	-	7471B	PPB	18 Apr 2025 09:20:40	U	SMPL
HighStd - 1	HighStd	50002	9.7938	-	7471B	PPB	18 Apr 2025 09:22:55	U	SMPL
ChkStd - 1	ChkStd	33255	6.5050	-	7471B	PPB	18 Apr 2025 09:25:11	U	SMPL
Q1821-01 - 1	SOIL-CUTTING	679	0.1077	-	7471B	PPB	18 Apr 2025 09:27:37	U	SMPL
Q1821-01DUP - 1	SOIL-CUTTINGDUP	523	0.0770	-	7471B	PPB	18 Apr 2025 09:32:38	U	SMPL
Q1821-01MS - 1	SOIL-CUTTINGMS	21558	4.2079	-	7471B	PPB	18 Apr 2025 09:34:54	U	SMPL
Q1821-01MSD - 1	SOIL-CUTTINGMSD	19444	3.7928	-	7471B	PPB	18 Apr 2025 09:40:14	U	SMPL
Q1821-01LX5 - 1		-145	-0.0541	-	7471B	PPB	18 Apr 2025 09:42:32	U	SMPL
Q1821-01A - 1		20421	3.9846	-	7471B	PPB	18 Apr 2025 09:44:52	U	SMPL
CCV93 - 1	CCV93	25187	4.9206	-	7471B	PPB	18 Apr 2025 09:50:59	U	SMPL
CCB93 - 1	CCB93	-39	-0.0333	-	7471B	PPB	18 Apr 2025 09:53:13	U	SMPL
Q1825-01DLX5 - 1	TP-9	14628	2.8470	-	7471B	PPB	18 Apr 2025 09:55:31	U	SMPL
Q1825-04DLX5 - 1	TP-10	30918	6.0461	-	7471B	PPB	18 Apr 2025 09:57:46	U	SMPL
Q1826-05 - 1	CONCRETE-FLOOR	-131	-0.0514	-	7471B	PPB	18 Apr 2025 10:00:04	U	SMPL
CCV94 - 1	CCV94	24943	4.8727	-	7471B	PPB	18 Apr 2025 10:02:30	U	SMPL
CCB94 - 1	CCB94	173	0.0083	-	7471B	PPB	18 Apr 2025 10:04:45	U	SMPL