

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135497

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/21/2025	09:05
2	CAL2	1	Water	NA	NA	20.2	7.00	04/21/2025	09:06
3	CAL3	1	Water	NA	NA	20.2	10.02	04/21/2025	09:09
4	ICV	1	Water	NA	NA	20.2	7.01	04/21/2025	09:11
5	CCV1	1	Water	NA	NA	20.2	2.01	04/21/2025	09:14
6	Q1841-03	1	Water	NA	NA	22.1	6.59	04/21/2025	09:25
7	Q1841-03DUP	1	Water	NA	NA	22.2	6.60	04/21/2025	09:26
8	CCV2	1	Water	NA	NA	20.1	12.02	04/21/2025	09:30

WORKLIST(Hardcopy Internal Chain)

VB 13549 f

WorkList Name : PH Q1841

WorkList ID : 189054

Department : Wet-Chemistry

Date : 04-21-2025 08:48:34

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1841-03	D OILY-WATER	Water	pH	Cool 4 deg C	PSEG03	L51	04/18/2025	9040C

Date/Time 04/21/25 09:00

Raw Sample Received by: JG WDC

Raw Sample Relinquished by: JG SM

Date/Time 04/21/25

Raw Sample Received by: JG SM

Raw Sample Relinquished by: JG WDC