

**Analytical Summary Report**

**Analysis Method:** 1010B

**Reviewed By:** rubina

**Parameter:** Flash Point

**Supervisor Review By:** Iwona

**Run Number:** LB135502

**Ambient Barometric Pressure (mmHg):** 771.00

**Thermometer ID:** Flash Point

**Barometric Scale ID:** 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.7	04/21/2025	08:20
2	Q1824-01		1	5	40.00	104.0	103.3	04/21/2025	08:50
3	Q1824-01DUP		1	5	40.00	104.0	103.3	04/21/2025	09:20
4	Q1824-02		1	6	40.00	104.0	103.3	04/21/2025	09:50

$$\text{Result} = (\text{Celsius} * 1.8) + 32$$

$$\text{Final Result} = \text{Result} + (760 - \text{Ambient Barometric Pressure}) * 0.06$$

WORKLIST(Hardcopy Internal Chain)

LB 135502

WorkList Name : fish point

WorkList ID : 189035

Department : Wet-Chemistry

Date : 04-21-2025 08:01:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1824-01	284	Solid	Flash Point	Cool 4 deg C	PSEG03	L31	04/16/2025	1010B
Q1824-02	286	Solid	Flash Point	Cool 4 deg C	PSEG03	L31	04/16/2025	1010B

Date/Time 04/21/2025 08:10  
 Raw Sample Received by: RM(wj)  
 Raw Sample Relinquished by: RM(wj)

Date/Time 04/21/2025 12:40  
 Raw Sample Received by: RM(wj)  
 Raw Sample Relinquished by: RM(wj)