

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB135503

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.7	04/21/2025	10:30
2	Q1824-04		1	5	43.00	109.4	108.7	04/21/2025	11:00
3	Q1841-03		1	15	100.00	>212.0	>212.0	04/21/2025	11:30
4	Q1841-03DUP		1	15	100.00	>212.0	>212.0	04/21/2025	12:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16135503

WorkList Name : flas point

WorkList ID : 189034

Department : Wet-Chemistry

Date : 04-21-2025 08:02:47

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1824-04	285	Water	Flash Point	Cool 4 deg C	PSEG03	L31	04/16/2025	1010B
Q1841-03	OILY-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	L51	04/18/2025	1010B

Date/Time 04/21/2025 08:10
 Raw Sample Received by: RM (w.c.)
 Raw Sample Relinquished by: JA (w.c.)

Date/Time 04/21/2025 12:40
 Raw Sample Received by: JA (w.c.)
 Raw Sample Relinquished by: RM (w.c.)