

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 4/22/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 04/21/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:25
Out Date: 04/22/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB135506

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q1837-15	FLORISIL CHECK	1	1.00	1.00	2.00	2.00	100.0	
Q1845-01	GAS-PIPE-1	2	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1846-01	JEI294J-1-1	12	1.00	1.00	2.00	2.00	100.0	pile
Q1846-02	JEI294J-1-2	13	1.00	1.00	2.00	2.00	100.0	pile
Q1847-01	MOO-25-0121	14	1.18	10.13	11.31	10.13	88.4	
Q1847-02	MOO-25-0122	15	1.15	9.96	11.11	10.58	94.7	
Q1847-03	PAD-24125	16	1.18	10.49	11.67	11.22	95.7	
Q1848-01	RBR200027	17	1.13	10.13	11.26	10.14	88.9	
Q1848-02	RBR200027-E2	18	1.15	10.37	11.52	9.94	84.8	
Q1848-03	ETGI-275	19	1.18	10.98	12.16	10.61	85.9	
Q1848-04	ETGI-275-E2	20	1.17	9.95	11.12	9.8	86.7	
Q1848-05	ETGI-342	21	1.14	10.56	11.7	10.38	87.5	
Q1848-06	ETGI-342-E2	22	1.16	10.33	11.49	10.15	87.0	
Q1850-01	CONCRETE-PILE-N-C-LOT	3	1.00	1.00	2.00	2.00	100.0	CONCRETE- PILE
Q1850-02	CONCRETE-PILE-N-C-LOT	4	1.00	1.00	2.00	2.00	100.0	CONCRETE- PILE
Q1851-01	DP5	5	1.14	9.88	11.02	10.7	96.8	
Q1851-02	DP6	6	1.14	9.88	11.02	10.72	97.0	
Q1851-03	DP7	7	1.15	9.82	10.97	10.82	98.5	
Q1851-04	DP8	8	1.13	10.84	11.97	11.14	92.3	
Q1851-05	GP5	9	1.14	10.60	11.74	11.47	97.5	
Q1851-06	GP6	10	1.13	11.10	12.23	11.00	88.9	
Q1851-07	GP7	11	1.18	10.44	11.62	10.91	93.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$