


LB135518

245.1

INSTRUMENT-10: CV2

 μ Abs.:

44812



Linear

A= 0.0000e+000

B= 2.2527e-004

C= 2.6054e-002

Rho= 0.9998042

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/1
0.0	0.000	0.031	0.031	23	0.000	23					-
0.05	0.050	0.060	0.010	150	0.0 %	150					20
0.2	0.200	0.259	0.059	1034	0.0 %	1034	0				30
2.5	2.500	2.466	-0.034	10832	0.0 %	10832					-1
5.0	5.000	4.908	-0.092	21673	0.0 %	21673					-2
7.5	7.500	7.405	-0.095	32755	0.0 %	32755					-1
10.0	10.000	10.121	0.121	44812	0.0 %	44812					1

LB135518 INSTRUMENT D : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 22 Apr 2025 13:33:21

Sample ID	Extended ID	Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	23	-	0.0000	245.1	PPB	22 Apr 2025 13:37:04	S	Std
0.05 - 1	50.5	150	-	0.0500	245.1	PPB	22 Apr 2025 13:42:03	S	Std
0.2 - 1	50.2	1034	-	0.2000	245.1	PPB	22 Apr 2025 13:44:20	S	Std
2.5 - 1	52.5	10832	-	2.5000	245.1	PPB	22 Apr 2025 13:46:37	S	Std
5.0 - 1	55	21673	-	5.0000	245.1	PPB	22 Apr 2025 13:48:54	S	Std
7.5 - 1	57.5	32755	-	7.5000	245.1	PPB	22 Apr 2025 13:51:11	S	Std
10.0 - 1	59.5	44812	-	10.0000	245.1	PPB	22 Apr 2025 13:57:09	S	Std
ICV97 - 1	ICV97	18499	4.1933	-245.1		PPB	22 Apr 2025 14:00:14	U	SMPL
ICB97 - 1	ICB97	-358	-0.0546	-245.1		PPB	22 Apr 2025 14:02:29	U	SMPL
CCV12 - 1	CCV12	22160	5.0180	-245.1		PPB	22 Apr 2025 14:04:48	U	SMPL
CCB12 - 1	CCB12	-368	-0.0568	-245.1		PPB	22 Apr 2025 14:07:02	U	SMPL
CRA - 1	CRA	90	0.0463	-245.1		PPB	22 Apr 2025 14:09:19	U	SMPL
HighStd - 1	HighStd	45542	10.2852	-245.1		PPB	22 Apr 2025 14:11:34	U	SMPL
ChkStd - 1	ChkStd	29279	6.6217	-245.1		PPB	22 Apr 2025 14:13:50	U	SMPL
PB167703BL - 1	PBW	-489	-0.0841	-245.1		PPB	22 Apr 2025 14:16:10	U	SMPL
PB167703BS - 1	LCSW	17839	4.0446	-245.1		PPB	22 Apr 2025 14:18:29	U	SMPL
Q1833-02 - 1	EFF-WASTE-WATER	-236	-0.0271	-245.1		PPB	22 Apr 2025 14:20:45	U	SMPL
Q1833-02DUP - 1	EFF-WASTE-WATERDUP	16	0.0297	-245.1		PPB	22 Apr 2025 14:23:03	U	SMPL
Q1833-02MS - 1	EFF-WASTE-WATERMS	15480	3.5132	-245.1		PPB	22 Apr 2025 14:29:43	U	SMPL
Q1833-02MSD - 1	EFF-WASTE-WATERMSD	16866	3.8254	-245.1		PPB	22 Apr 2025 14:32:01	U	SMPL
CCV13 - 1	CCV13	22064	4.9964	-245.1		PPB	22 Apr 2025 14:34:19	U	SMPL
CCB13 - 1	CCB13	-373	-0.0580	-245.1		PPB	22 Apr 2025 14:36:35	U	SMPL
Q1833-02LX5 - 1		-63	0.0119	-245.1		PPB	22 Apr 2025 14:38:53	U	SMPL
Q1833-02A - 1		16723	3.7932	-245.1		PPB	22 Apr 2025 14:41:08	U	SMPL
CCV14 - 1	CCV14	22047	4.9925	-245.1		PPB	22 Apr 2025 14:43:24	U	SMPL
CCB14 - 1	CCB14	-339	-0.0503	-245.1		PPB	22 Apr 2025 14:45:41	U	SMPL