

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135525

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/23/2025	08:22
2	CAL2	1	Water	NA	NA	20.2	7.01	04/23/2025	08:25
3	CAL3	1	Water	NA	NA	20.2	10.02	04/23/2025	08:26
4	ICV	1	Water	NA	NA	20.2	7.01	04/23/2025	08:30
5	CCV1	1	Water	NA	NA	20.2	2.01	04/23/2025	08:33
6	Q1855-06	1	Water	NA	NA	24.2	5.30	04/23/2025	08:40
7	Q1855-06DUP	1	Water	NA	NA	24.3	5.31	04/23/2025	08:41
8	CCV2	1	Water	NA	NA	20.2	12.02	04/23/2025	08:44

WORKLIST(Hardcopy Internal Chain)

17135525

WorkList Name : PH Q W1855

WorkList ID : 189090

Department : Wet-Chemistry

Date : 04-23-2025 08:04:39

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1855-06	A 60452	Water	pH	Cool 4 deg C	PSEG03	L51	04/22/2025	9040C

6/11/2025 08:15

Date/Time
Raw Sample Received by: AA WDC
Raw Sample Relinquished by: JD CSM

Date/Time
Raw Sample Received by: 04/29/25
Raw Sample Relinquished by: SA WDC