

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135526

Slope : 99.2

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/23/2025	08:22
2	CAL2	1	Water	NA	NA	20.2	7.01	04/23/2025	08:25
3	CAL3	1	Water	NA	NA	20.2	10.02	04/23/2025	08:26
4	ICV	1	Water	NA	NA	20.2	7.01	04/23/2025	08:30
5	CCV1	1	Water	NA	NA	20.2	2.02	04/23/2025	08:33
6	Q1855-05	1	Solid	20.02	20	23.7	9.61	04/23/2025	09:10
7	Q1855-05DUP	1	Solid	20.03	20	23.8	9.62	04/23/2025	09:11
8	CCV2	1	Water	NA	NA	20.3	12.02	04/23/2025	09:14

WORKLIST(Hardcopy Internal Chain)

135526

WorkList Name : PH S Q1855 WorkList ID : 189089 Department : Wet-Chemistry Date : 04-23-2025 08:03:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1855-05	C 60444	Solid	pH	Cool 4 deg C	PSEG03	L51	04/22/2025	9045D

Date/Time 04/23/25 08:15
Raw Sample Received by: JF Wood
Raw Sample Relinquished by: JD CSM

Date/Time 04/23/25 12:30
Raw Sample Received by: JD CSM
Raw Sample Relinquished by: JF Wood