

7470A

INSTRUMENT ID: CV1

Linear

$$\mu \text{ Abs.} :$$

45191

A= 0.0000e+000

B= 2.2129e-004

C= -1.1005e-002

Rho= 0.9999010

Accept=Accepted

Max. Conc:— 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.028	-0.028	-78	0.000	-78	0				-
0.2	0.200	0.215	0.015	1023	0.0 %	1023					7
2.5	2.500	2.464	-0.036	11186	0.0 %	11186					-1
5.0	5.000	5.107	0.107	23128	0.0 %	23128					2
7.5	7.500	7.452	-0.048	33725	0.0 %	33725					-1
10.0	10.000	9.989	-0.011	45191	0.0 %	45191					0

LB135529 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Apr 2025 11:23:17

Sample ID	Extended ID	μ Abs	Conc	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-78	-	0.00007470A	PPB	23 Apr 2025 11:46:47	S	Std
0.2 - 1	50-2	1023	-	0.20007470A	PPB	23 Apr 2025 11:49:04	S	Std
2.5 - 1	50-3	11186	-	2.50007470A	PPB	23 Apr 2025 11:51:21	S	Std
5.0 - 1	50-4	23128	-	5.00007470A	PPB	23 Apr 2025 11:53:37	S	Std
7.5 - 1	50-5	33725	-	7.50007470A	PPB	23 Apr 2025 11:55:55	S	Std
10.0 - 1	50-6	45191	-	10.00007470A	PPB	23 Apr 2025 11:58:13	S	Std
ICV98 - 1	ICV98	16635	3.6702	-7470A	PPB	23 Apr 2025 12:02:02	U	SMPL
ICB98 - 1	ICB98	-257	-0.0679	-7470A	PPB	23 Apr 2025 12:04:17	U	SMPL
CCV14 - 1	CCV14	22825	5.0400	-7470A	PPB	23 Apr 2025 12:06:34	U	SMPL
CCB14 - 1	CCB14	-215	-0.0586	-7470A	PPB	23 Apr 2025 12:08:49	U	SMPL
CRA - 1	CRA	999	0.2101	-7470A	PPB	23 Apr 2025 12:11:06	U	SMPL
HighStd - 1	HighStd	44839	9.9115	-7470A	PPB	23 Apr 2025 12:13:22	U	SMPL
ChkStd - 1	ChkStd	28580	6.3135	-7470A	PPB	23 Apr 2025 12:15:38	U	SMPL
PB167714BL - 1	PBW	-459	-0.1126	-7470A	PPB	23 Apr 2025 12:17:57	U	SMPL
PB167714BS - 1	LCSW	16780	3.7023	-7470A	PPB	23 Apr 2025 12:20:15	U	SMPL
Q1842-06 - 1	PL-714-COMP-08	-313	-0.0803	-7470A	PPB	23 Apr 2025 12:22:31	U	SMPL
Q1842-12 - 1	PL-714-COMP-09	-38	-0.0194	-7470A	PPB	23 Apr 2025 12:24:49	U	SMPL
Q1842-18 - 1	PL-714-COMP-10	8	-0.0092	-7470A	PPB	23 Apr 2025 12:27:07	U	SMPL
Q1842-24 - 1	PL-714-COMP-11	39	-0.0024	-7470A	PPB	23 Apr 2025 12:29:23	U	SMPL
Q1842-30 - 1	PL-714-COMP-12	27	-0.0050	-7470A	PPB	23 Apr 2025 12:31:39	U	SMPL
CCV15 - 1	CCV15	22850	5.0455	-7470A	PPB	23 Apr 2025 12:33:55	U	SMPL
CCB15 - 1	CCB15	-246	-0.0654	-7470A	PPB	23 Apr 2025 12:36:10	U	SMPL
Q1842-36 - 1	PL-714-COMP-13	7	-0.0095	-7470A	PPB	23 Apr 2025 12:38:27	U	SMPL
Q1842-42 - 1	PL-714-COMP-14	38	-0.0026	-7470A	PPB	23 Apr 2025 12:40:42	U	SMPL
Q1850-02 - 1	CONCRETE-PILE-N-C-LOT	17	-0.0072	-7470A	PPB	23 Apr 2025 12:42:58	U	SMPL
Q1850-02DUP - 1	CONCRETE-PILE-N-C-LOTDUP	31	-0.0041	-7470A	PPB	23 Apr 2025 12:45:13	U	SMPL
Q1850-02MS - 1	CONCRETE-PILE-N-C-LOTMS	16453	3.6299	-7470A	PPB	23 Apr 2025 12:52:33	U	SMPL
Q1850-02MSD - 1	CONCRETE-PILE-N-C-LOTMSD	16285	3.5927	-7470A	PPB	23 Apr 2025 12:54:49	U	SMPL
PB167660TB - 1	PB167660TB	-258	-0.0681	-7470A	PPB	23 Apr 2025 12:57:07	U	SMPL
Q1850-02LX5 - 1		-5	-0.0121	-7470A	PPB	23 Apr 2025 12:59:26	U	SMPL
Q1850-02A - 1		16462	3.6319	-7470A	PPB	23 Apr 2025 13:01:43	U	SMPL
CCV16 - 1	CCV16	22422	4.9508	-7470A	PPB	23 Apr 2025 13:03:59	U	SMPL
CCB16 - 1	CCB16	-326	-0.0831	-7470A	PPB	23 Apr 2025 13:06:16	U	SMPL