

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB135537

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/23/2025	15:20
2	CAL2	1	Water	NA	NA	20.2	7.00	04/23/2025	15:21
3	CAL3	1	Water	NA	NA	20.3	10.02	04/23/2025	15:22
4	ICV	1	Water	NA	NA	20.2	7.00	04/23/2025	15:25
5	CCV1	1	Water	NA	NA	20.2	2.01	04/23/2025	15:30
6	Q1852-02	1	Solid	20.02	20	24.6	7.61	04/23/2025	15:37
7	Q1852-02DUP	1	Solid	20.03	20	24.7	7.62	04/23/2025	15:38
8	Q1852-04	1	Solid	20.02	20	24.5	8.26	04/23/2025	15:42
9	Q1852-06	1	Solid	20.04	20	24.3	8.16	04/23/2025	15:45
10	Q1852-08	1	Solid	20.02	20	24.0	8.01	04/23/2025	15:55
11	Q1855-04	1	Solid	20.03	20	23.9	9.27	04/23/2025	16:10
12	Q1865-01	1	Solid	20.02	20	23.8	6.52	04/23/2025	16:15
13	Q1865-02	1	Solid	20.02	20	23.6	7.45	04/23/2025	16:20
14	Q1865-03	1	Solid	20.03	20	23.9	7.76	04/23/2025	16:25
15	Q1865-04	1	Solid	20.02	20	24.2	7.80	04/23/2025	16:30
16	CCV2	1	Water	NA	NA	20.3	12.02	04/23/2025	16:35

WORKLIST(Hardcopy Internal Chain)

UR 135338

WorkList Name : corrsoivity q1752 WorkList ID : 189104 Department : Wet-Chemistry Date : 04-23-2025 12:47:08

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1852-02	ETGI-354	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	04/22/2025	9045D
Q1852-04	72-11977	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	04/22/2025	9045D
Q1852-06	ETGI-278	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	04/22/2025	9045D
Q1852-08	72-12013	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	04/22/2025	9045D
Q1855-04	2001-2002	Solid	Corrosivity	Cool 4 deg C	PSEG03	L51	04/22/2025	9045D
Q1865-01	COMP-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	04/23/2025	9045D
Q1865-02	COMP-2	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	04/23/2025	9045D
Q1865-03	COMP-3	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	04/23/2025	9045D
Q1865-04	COMP-4	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	04/23/2025	9045D

Date/Time 04/23/25 15:10

Raw Sample Received by: SCW

Raw Sample Relinquished by: SCW

Date/Time 04/23/25 17:30

Raw Sample Received by: SCW

Raw Sample Relinquished by: SCW