

LB135549

7470A

INSTRUMENT IN CV2

Linear

 μ Abs.:

50789

A= 0.0000e+000

B= 1.9890e-004

C= -5.5341e-003

Rho= 0.9994639

Accept=Accepted

Max. Conc: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.008	0.008	68	0.000	68	0				1
0.2	0.200	0.202	0.002	1042	0.0 %	1042					1
2.5	2.500	2.475	-0.025	12472	0.0 %	12472					1
5.0	5.000	5.152	0.152	25931	0.0 %	25931					1
7.5	7.500	7.267	-0.233	36562	0.0 %	36562					3
10.0	10.000	10.096	0.096	50789	0.0 %	50789					3

LB135547 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Apr 2025 12:46:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	68	-	0.0000	7470A	PPB	24 Apr 2025 12:55:11	S	Std
0.2 - 1	50-2	1042	-	0.2000	7470A	PPB	24 Apr 2025 12:57:28	S	Std
2.5 - 1	52.5	12472	-	2.5000	7470A	PPB	24 Apr 2025 12:59:46	S	Std
5.0 - 1	55	25931	-	5.0000	7470A	PPB	24 Apr 2025 13:02:03	S	Std
7.5 - 1	57.5	36562	-	7.5000	7470A	PPB	24 Apr 2025 13:04:21	S	Std
10.0 - 1	51	50789	-	10.0000	7470A	PPB	24 Apr 2025 13:11:09	S	Std
ICV100 - 1	ICV100	19818	3.9363	-	7470A	PPB	24 Apr 2025 13:16:58	U	SMPL
ICB100 - 1	ICB100	-174	-0.0401	-	7470A	PPB	24 Apr 2025 13:19:13	U	SMPL
CCV21 - 1	CCV21	23283	4.6254	-	7470A	PPB	24 Apr 2025 13:21:30	U	SMPL
CCB21 - 1	CCB21	-219	-0.0491	-	7470A	PPB	24 Apr 2025 13:23:45	U	SMPL
CRA - 1	CRA	965	0.1864	-	7470A	PPB	24 Apr 2025 13:26:02	U	SMPL
HighStd - 1	HighStd	46485	9.2403	-	7470A	PPB	24 Apr 2025 13:31:08	U	SMPL
ChkStd - 1	ChkStd	36059	7.1666	-	7470A	PPB	24 Apr 2025 13:36:07	U	SMPL
PB167724BL - 1	PBW	-79	-0.0212	-	7470A	PPB	24 Apr 2025 13:41:05	U	SMPL
PB167724BS - 1	LCSW	19017	3.7769	-	7470A	PPB	24 Apr 2025 13:45:55	U	SMPL
Q1791-15 - 1	279312	-244	-0.0541	-	7470A	PPB	24 Apr 2025 13:48:13	U	SMPL
Q1791-16 - 1	290502	-48	-0.0151	-	7470A	PPB	24 Apr 2025 13:50:31	U	SMPL
Q1852-02 - 1	ETGI-354	-39	-0.0133	-	7470A	PPB	24 Apr 2025 13:52:47	U	SMPL
Q1852-04 - 1	72-11977	-14	-0.0083	-	7470A	PPB	24 Apr 2025 13:55:04	U	SMPL
Q1852-06 - 1	ETGI-278	7	-0.0041	-	7470A	PPB	24 Apr 2025 13:57:20	U	SMPL
CCV22 - 1	CCV22	26619	5.2890	-	7470A	PPB	24 Apr 2025 14:04:19	U	SMPL
CCB22 - 1	CCB22	-243	-0.0539	-	7470A	PPB	24 Apr 2025 14:06:34	U	SMPL
Q1852-08 - 1	72-12013	-20	-0.0095	-	7470A	PPB	24 Apr 2025 14:08:51	U	SMPL
Q1855-04 - 1	2001-2002	85	0.0114	-	7470A	PPB	24 Apr 2025 14:11:07	U	SMPL
Q1865-01 - 1		-7	-0.0069	-	7470A	PPB	24 Apr 2025 14:13:22	U	SMPL
Q1865-02 - 1		9	-0.0037	-	7470A	PPB	24 Apr 2025 14:15:38	U	SMPL
Q1865-03 - 1		-4	-0.0063	-	7470A	PPB	24 Apr 2025 14:17:54	U	SMPL
PB167725BL - 1	PBW	12	-0.0031	-	7470A	PPB	24 Apr 2025 14:20:10	U	SMPL
PB167725BS - 1	LCSW	19626	3.8981	-	7470A	PPB	24 Apr 2025 14:25:39	U	SMPL
Q1791-17 - 1	VB16193	-191	-0.0435	-	7470A	PPB	24 Apr 2025 14:27:55	U	SMPL
Q1791-17DUP - 1	VB16193DUP	-28	-0.0111	-	7470A	PPB	24 Apr 2025 14:30:15	U	SMPL
Q1791-17MS - 1	VB16193MS	19352	3.8436	-	7470A	PPB	24 Apr 2025 14:32:31	U	SMPL
CCV23 - 1	CCV23	22804	4.5302	-	7470A	PPB	24 Apr 2025 14:34:46	U	SMPL
CCB23 - 1	CCB23	-265	-0.0582	-	7470A	PPB	24 Apr 2025 14:37:03	U	SMPL
Q1791-17MSD - 1	VB16193MSD	18701	3.7141	-	7470A	PPB	24 Apr 2025 14:39:20	U	SMPL
Q1791-18 - 1	280722	-212	-0.0477	-	7470A	PPB	24 Apr 2025 14:41:36	U	SMPL
Q1791-19 - 1	280791	-13	-0.0081	-	7470A	PPB	24 Apr 2025 14:43:54	U	SMPL
Q1791-20 - 1	V1207	15	-0.0026	-	7470A	PPB	24 Apr 2025 14:46:10	U	SMPL
Q1791-21 - 1	VB16164	3	-0.0049	-	7470A	PPB	24 Apr 2025 14:48:26	U	SMPL
Q1865-04 - 1		23	-0.0010	-	7470A	PPB	24 Apr 2025 14:50:43	U	SMPL
Q1865-04DUP - 1		1	-0.0053	-	7470A	PPB	24 Apr 2025 14:52:59	U	SMPL
Q1865-04MS - 1		19557	3.8843	-	7470A	PPB	24 Apr 2025 14:55:16	U	SMPL
Q1865-04MSD - 1		18461	3.6664	-	7470A	PPB	24 Apr 2025 14:59:31	U	SMPL
CCV24 - 1	CCV24	22949	4.5590	-	7470A	PPB	24 Apr 2025 15:04:28	U	SMPL
CCB24 - 1	CCB24	-280	-0.0612	-	7470A	PPB	24 Apr 2025 15:06:43	U	SMPL
PB167706TB - 1	PB167706TB	-32	-0.0119	-	7470A	PPB	24 Apr 2025 15:09:01	U	SMPL
PB167717TB - 1	PB167717TB	-14	-0.0083	-	7470A	PPB	24 Apr 2025 15:11:16	U	SMPL
Q1865-04LX5 - 1		19	-0.0018	-	7470A	PPB	24 Apr 2025 15:13:32	U	SMPL
Q1865-04A - 1		17178	3.4112	-	7470A	PPB	24 Apr 2025 15:15:47	U	SMPL
Q1791-17LX5 - 1		-529	-0.1108	-	7470A	PPB	24 Apr 2025 15:18:03	U	SMPL
Q1791-17A - 1		13641	2.7077	-	7470A	PPB	24 Apr 2025 15:25:47	U	SMPL
CCV25 - 1	CCV25	26939	5.3526	-	7470A	PPB	24 Apr 2025 15:28:05	U	SMPL
CCB25 - 1	CCB25	-263	-0.0578	-	7470A	PPB	24 Apr 2025 15:30:23	U	SMPL