

LA13551


74719
7470A

INSTRUMENT ID: CY1

Linear

μ Abs.:

46101



A= 0.0000e+000

B= 2.2075e-004

C= 1.9937e-002

Rho= 0.9993043

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	40D
0.0	0.000	0.025	0.025	21	0.000	21	0				40D
0.2	0.200	0.238	0.038	986	0.0 %	986					19
2.5	2.500	2.569	0.069	11545	0.0 %	11545					3
5.0	5.000	4.922	-0.078	22205	0.0 %	22205					2
7.5	7.500	7.251	-0.249	32756	0.0 %	32756					2
10.0	10.000	10.197	0.197	46101	0.0 %	46101					2

LB135551 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 25 Apr 2025 10:07:51

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	21	-	0.0000	7470A	PPB	25 Apr 2025 10:14:45	S	Std
0.2 - 1	50.2	986	-	0.2000	7470A	PPB	25 Apr 2025 10:17:02	S	Std
2.5 - 1	52.5	11545	-	2.5000	7470A	PPB	25 Apr 2025 10:19:19	S	Std
5.0 - 1	55	22205	-	5.0000	7470A	PPB	25 Apr 2025 10:21:36	S	Std
7.5 - 1	57.5	32756	-	7.5000	7470A	PPB	25 Apr 2025 10:23:54	S	Std
10.0 - 1	110	46101	-	10.0000	7470A	PPB	25 Apr 2025 10:26:12	S	Std
ICV101 - 1	ICV101	16340	3.6270	-	7470A	PPB	25 Apr 2025 10:29:13	U	SMPL
ICB101 - 1	ICB101	-177	-0.0191	-	7470A	PPB	25 Apr 2025 10:31:28	U	SMPL
CCV25 - 1	CCV25	21585	4.7848	-	7470A	PPB	25 Apr 2025 10:33:45	U	SMPL
CCB25 - 1	CCB25	-225	-0.0297	-	7470A	PPB	25 Apr 2025 10:36:00	U	SMPL
CRA - 1	CRA	842	0.2058	-	7470A	PPB	25 Apr 2025 10:38:18	U	SMPL
HighStd - 1	HighStd	45344	10.0297	-	7470A	PPB	25 Apr 2025 10:40:33	U	SMPL
ChkStd - 1	ChkStd	28008	6.2027	-	7470A	PPB	25 Apr 2025 10:42:49	U	SMPL
PB167745BL - 1	PBS	-310	-0.0485	-	7470A	PPB	25 Apr 2025 10:45:07	U	SMPL
PB167745BS - 1	LCSS	15360	3.4107	-	7470A	PPB	25 Apr 2025 10:47:26	U	SMPL
Q1867-01 - 1	RBR200059	2295	0.5266	-	7470A	PPB	25 Apr 2025 10:49:42	U	SMPL
Q1867-01DUP - 1	RBR200059DUP	1983	0.4577	-	7470A	PPB	25 Apr 2025 10:52:01	U	SMPL
Q1867-01MS - 1	RBR200059MS	18143	4.0250	-	7470A	PPB	25 Apr 2025 10:54:19	U	SMPL
Q1867-01MSD - 1	RBR200059MSD	17407	3.8625	-	7470A	PPB	25 Apr 2025 10:56:37	U	SMPL
Q1867-03 - 1	VNJ211	4804	1.0804	-	7470A	PPB	25 Apr 2025 10:58:55	U	SMPL
CCV26 - 1	CCV26	20890	4.6314	-	7470A	PPB	25 Apr 2025 11:01:13	U	SMPL
CCB26 - 1	CCB26	-114	-0.0052	-	7470A	PPB	25 Apr 2025 11:03:31	U	SMPL
Q1869-01 - 1	MH-F	1513	0.3539	-	7470A	PPB	25 Apr 2025 11:05:49	U	SMPL
Q1871-01 - 1	MH-A	1456	0.3414	-	7470A	PPB	25 Apr 2025 11:08:07	U	SMPL
Q1871-05 - 1	MH-B	524	0.1356	-	7470A	PPB	25 Apr 2025 11:10:25	U	SMPL
Q1874-01 - 1	VNJ-236	769	0.1897	-	7470A	PPB	25 Apr 2025 11:12:43	U	SMPL
Q1874-03 - 1	RT1491	857	0.2091	-	7470A	PPB	25 Apr 2025 11:15:01	U	SMPL
Q1874-05 - 1	HT3727	1174	0.2791	-	7470A	PPB	25 Apr 2025 11:17:19	U	SMPL
Q1875-01 - 1	AUD-25-0053	905	0.2197	-	7470A	PPB	25 Apr 2025 11:19:37	U	SMPL
Q1877-01 - 1	AU-6-042425	799	0.1963	-	7470A	PPB	25 Apr 2025 11:21:55	U	SMPL
Q1878-01 - 1	TR-4-042425	833	0.2038	-	7470A	PPB	25 Apr 2025 11:24:13	U	SMPL
CCV27 - 1	CCV27	21115	4.6811	-	7470A	PPB	25 Apr 2025 11:26:31	U	SMPL
CCB27 - 1	CCB27	-127	-0.0081	-	7470A	PPB	25 Apr 2025 11:28:49	U	SMPL
Q1867-01LX5 - 1		562	0.1440	-	7470A	PPB	25 Apr 2025 11:31:07	U	SMPL
Q1867-01A - 1		19231	4.2652	-	7470A	PPB	25 Apr 2025 11:33:25	U	SMPL
CCV28 - 1	CCV28	20870	4.6270	-	7470A	PPB	25 Apr 2025 11:35:43	U	SMPL
CCB28 - 1	CCB28	-169	-0.0174	-	7470A	PPB	25 Apr 2025 11:38:01	U	SMPL