

7471B

INSTRUMENT 11: CV1

Linear

$$\mu \text{ Abs.} :$$

37294

A= 0.0000e+000

B= 2.6476e-004

C= -7.2388e-002

Rho= 0.9986611

Accept=Accepted

Max. Conc:— 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/10
0.0	0.000	-0.049	-0.049	88	0.000	88					
0.2	0.200	0.184	-0.016	969	0.0 %	969					
2.5	2.500	2.346	-0.154	9135	0.0 %	9135					
5.0	5.000	5.390	0.390	20632	0.0 %	20632					
7.5	7.500	7.527	0.027	28703	0.0 %	28703					
10.0	10.000	9.802	-0.198	37294	0.0 %	37294					

LB135936 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 28 May 2025 14:53:44

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	88	-	0.0000	7471B	PPB	28 May 2025 15:56:18	S	Std
0.2 - 1	50.2	969	-	0.2000	7471B	PPB	28 May 2025 15:58:36	S	Std
2.5 - 1	52.5	9135	-	2.5000	7471B	PPB	28 May 2025 16:00:53	S	Std
5.0 - 1	55	20632	-	5.0000	7471B	PPB	28 May 2025 16:10:01	S	Std
7.5 - 1	57.5	28703	-	7.5000	7471B	PPB	28 May 2025 16:17:35	S	Std
10.0 - 1	510	37294	-	10.0000	7471B	PPB	28 May 2025 16:31:46	S	Std
ICV20 - 1	ICV20	15413	4.0084	-	7471B	PPB	28 May 2025 16:37:13	U	SMPL
ICB20 - 1	ICB20	80	-0.0512	-	7471B	PPB	28 May 2025 16:39:29	U	SMPL
CCV74 - 1	CCV74	21039	5.4979	-	7471B	PPB	28 May 2025 16:41:47	U	SMPL
CCB74 - 1	CCB74	-58	-0.0877	-	7471B	PPB	28 May 2025 16:44:03	U	SMPL
CRA - 1	CRA	1045	0.2043	-	7471B	PPB	28 May 2025 16:49:05	U	SMPL
HighStd - 1	HighStd	37195	9.7754	-	7471B	PPB	28 May 2025 16:51:20	U	SMPL
ChkStd - 1	ChkStd	26100	6.8378	-	7471B	PPB	28 May 2025 16:53:35	U	SMPL
PB168187BL - 1	PBS	242	-0.0083	-	7471B	PPB	28 May 2025 16:58:41	U	SMPL
PB168187BS - 1	LCSS	17429	4.5421	-	7471B	PPB	28 May 2025 17:00:56	U	SMPL
Q2128-01 - 1	TP03-MHM	1515	0.3287	-	7471B	PPB	28 May 2025 17:08:19	U	SMPL
Q2128-01DUP - 1	TP03-MHMDUP	1323	0.2779	-	7471B	PPB	28 May 2025 17:10:35	U	SMPL
Q2128-01MS - 1	TP03-MHMMS	17091	4.4526	-	7471B	PPB	28 May 2025 17:15:34	U	SMPL
Q2128-01MSD - 1	TP03-MHMMSD	17365	4.5252	-	7471B	PPB	28 May 2025 17:17:51	U	SMPL
Q2130-01 - 1	TP-3	762	0.1294	-	7471B	PPB	28 May 2025 17:20:09	U	SMPL
CCV75 - 1	CCV75	19197	5.0102	-	7471B	PPB	28 May 2025 17:22:27	U	SMPL
CCB75 - 1	CCB75	132	-0.0374	-	7471B	PPB	28 May 2025 17:24:42	U	SMPL
Q2128-01LX5 - 1		513	0.0634	-	7471B	PPB	28 May 2025 17:41:08	U	SMPL
Q2128-01A - 1		16271	4.2355	-	7471B	PPB	28 May 2025 17:43:24	U	SMPL
CCV76 - 1	CCV76	19035	4.9673	-	7471B	PPB	28 May 2025 17:45:41	U	SMPL
CCB76 - 1	CCB76	-67	-0.0901	-	7471B	PPB	28 May 2025 17:48:01	U	SMPL