

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB135572

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/28/2025	10:40
2	CAL2	1	Water	NA	NA	20.2	7.01	04/28/2025	10:41
3	CAL3	1	Water	NA	NA	20.3	10.02	04/28/2025	10:42
4	ICV	1	Water	NA	NA	20.3	7.01	04/28/2025	10:45
5	CCV1	1	Water	NA	NA	20.2	2.01	04/28/2025	10:46
6	Q1896-03	1	Water	NA	NA	24.2	7.11	04/28/2025	10:55
7	Q1896-03DUP	1	Water	NA	NA	24.3	7.12	04/28/2025	10:56
8	Q1897-01	1	Water	NA	NA	21.7	4.74	04/28/2025	11:00
9	Q1897-02	1	Water	NA	NA	22.3	5.43	04/28/2025	11:11
10	CCV2	1	Water	NA	NA	20.2	12.02	04/28/2025	11:15

WORKLIST(Hardcopy Internal Chain)

WB 135512

WorkList Name : ph q1896

WorkList ID : 189182

Department : Wet-Chemistry

Date : 04-28-2025 10:21:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1896-03	C 295-BERGEN-FRAC	Water	pH	Cool 4 deg C	PSEG03	L31	04/25/2025	9040C
Q1897-01	B 40125	Water	pH	Cool 4 deg C	PSEG03	L41	04/25/2025	9040C
Q1897-02	B 41425	Water	pH	Cool 4 deg C	PSEG03	L41	04/25/2025	9040C

Date/Time 04/28/25 10:30

Raw Sample Received by: SO WEC

Raw Sample Relinquished by: AL SM

Date/Time 04/28/25

Raw Sample Received by: AL SM

Raw Sample Relinquished by: SO WEC

13:30

AL SM

SO WEC