

LB135584


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

50241



A= 0.0000e+000

B= 2.0252e-004

C= 1.8357e-002

Rho= 0.9992531

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.015	0.015	-16	0.000	-16	0				0.00
0.2	0.200	0.234	0.034	1066	0.0 %	1066					17
2.5	2.500	2.621	0.121	12850	0.0 %	12850					5
5.0	5.000	4.855	-0.145	23883	0.0 %	23883					-3
7.5	7.500	7.282	-0.218	35864	0.0 %	35864					-3
10.0	10.000	10.193	0.193	50241	0.0 %	50241					2

LB135584 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 29 Apr 2025 08:30:02

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-16	-	0.0000	7470A	PPB	29 Apr 2025 08:30:15	S	Std
0.2 - 1	50.2	1066	-	0.2000	7470A	PPB	29 Apr 2025 08:32:32	S	Std
2.5 - 1	52.5	12850	-	2.5000	7470A	PPB	29 Apr 2025 08:34:48	S	Std
5.0 - 1	55	23883	-	5.0000	7470A	PPB	29 Apr 2025 08:37:06	S	Std
7.5 - 1	57.5	35864	-	7.5000	7470A	PPB	29 Apr 2025 08:39:23	S	Std
10.0 - 1	510	50241	-	10.0000	7470A	PPB	29 Apr 2025 08:41:41	S	Std
ICV104 - 1	ICV104	18288	3.7221	-	7470A	PPB	29 Apr 2025 08:44:45	U	SMPL
ICB104 - 1	ICB104	-196	-0.0213	-	7470A	PPB	29 Apr 2025 08:47:00	U	SMPL
CCV36 - 1	CCV36	23810	4.8404	-	7470A	PPB	29 Apr 2025 08:49:17	U	SMPL
CCB36 - 1	CCB36	-261	-0.0345	-	7470A	PPB	29 Apr 2025 08:51:33	U	SMPL
CRA - 1	CRA	1035	0.2280	-	7470A	PPB	29 Apr 2025 08:53:50	U	SMPL
HighStd - 1	HighStd	51086	10.3643	-	7470A	PPB	29 Apr 2025 08:56:05	U	SMPL
ChkStd - 1	ChkStd	30039	6.1019	-	7470A	PPB	29 Apr 2025 08:58:20	U	SMPL
PB167772BL - 1	PBW	-428	-0.0683	-	7470A	PPB	29 Apr 2025 09:00:39	U	SMPL
PB167772BS - 1	LCSW	21262	4.3243	-	7470A	PPB	29 Apr 2025 09:02:58	U	SMPL
Q1869-04 - 1	MH-F	-361	-0.0548	-	7470A	PPB	29 Apr 2025 09:05:14	U	SMPL
Q1871-04 - 1	MH-A	-14	0.0155	-	7470A	PPB	29 Apr 2025 09:07:32	U	SMPL
Q1871-08 - 1	MH-B	18	0.0220	-	7470A	PPB	29 Apr 2025 09:09:50	U	SMPL
Q1874-02 - 1	VNJ-236	145	0.0477	-	7470A	PPB	29 Apr 2025 09:12:06	U	SMPL
Q1874-04 - 1	RT1491	-13	0.0157	-	7470A	PPB	29 Apr 2025 09:14:22	U	SMPL
CCV37 - 1	CCV37	24371	4.9540	-	7470A	PPB	29 Apr 2025 09:16:37	U	SMPL
CCB37 - 1	CCB37	-279	-0.0381	-	7470A	PPB	29 Apr 2025 09:18:53	U	SMPL
Q1874-06 - 1	HT3727	-24	0.0135	-	7470A	PPB	29 Apr 2025 09:21:10	U	SMPL
Q1875-02 - 1	AUD-25-0053	3	0.0190	-	7470A	PPB	29 Apr 2025 09:23:25	U	SMPL
Q1891-04 - 1	MH-C	41	0.0267	-	7470A	PPB	29 Apr 2025 09:25:41	U	SMPL
Q1891-08 - 1	MH-D	36	0.0256	-	7470A	PPB	29 Apr 2025 09:27:57	U	SMPL
Q1892-04 - 1	MH-G	37	0.0259	-	7470A	PPB	29 Apr 2025 09:30:12	U	SMPL
Q1892-08 - 1	MH-H	11	0.0206	-	7470A	PPB	29 Apr 2025 09:32:29	U	SMPL
Q1892-12 - 1	MH-U2	15	0.0214	-	7470A	PPB	29 Apr 2025 09:34:45	U	SMPL
Q1895-02 - 1	COMP-1	-23	0.0137	-	7470A	PPB	29 Apr 2025 09:37:02	U	SMPL
Q1895-04 - 1	COMP-2	-3	0.0177	-	7470A	PPB	29 Apr 2025 09:39:20	U	SMPL
Q1895-06 - 1	COMP-3	-10	0.0163	-	7470A	PPB	29 Apr 2025 09:41:35	U	SMPL
CCV38 - 1	CCV38	24505	4.9811	-	7470A	PPB	29 Apr 2025 09:43:51	U	SMPL
CCB38 - 1	CCB38	-205	-0.0232	-	7470A	PPB	29 Apr 2025 09:46:07	U	SMPL
Q1900-04 - 1	WC-1	33	0.0250	-	7470A	PPB	29 Apr 2025 09:48:25	U	SMPL
Q1900-08 - 1	WC-2	35	0.0254	-	7470A	PPB	29 Apr 2025 09:50:40	U	SMPL
Q1900-12 - 1	WC-3	44	0.0273	-	7470A	PPB	29 Apr 2025 09:52:56	U	SMPL
Q1900-12DUP - 1	WC-3DUP	47	0.0279	-	7470A	PPB	29 Apr 2025 09:55:11	U	SMPL
Q1900-12MS - 1	WC-3MS	17663	3.5955	-	7470A	PPB	29 Apr 2025 09:57:27	U	SMPL
Q1900-12MSD - 1	WC-3MSD	17693	3.6016	-	7470A	PPB	29 Apr 2025 10:02:23	U	SMPL
PB167743TB - 1	PB167743TB	-216	-0.0254	-	7470A	PPB	29 Apr 2025 10:04:40	U	SMPL
Q1900-12LX5 - 1		11	0.0206	-	7470A	PPB	29 Apr 2025 10:06:59	U	SMPL
Q1900-12A - 1		17448	3.5519	-	7470A	PPB	29 Apr 2025 10:09:16	U	SMPL
CCV39 - 1	CCV39	25287	5.1395	-	7470A	PPB	29 Apr 2025 10:11:31	U	SMPL
CCB39 - 1	CCB39	-165	-0.0151	-	7470A	PPB	29 Apr 2025 10:13:49	U	SMPL